

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101121\
 Data File : PD066156.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Oct 2021 10:30
 Operator : AR\AJ
 Sample : M4117-17DL 30X
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 SB-5(3.0-3.5)DL

Manual Integrations
 APPROVED

mohammad
 10/13/2021 1:19:16 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 12 01:38:22 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD092221.M
 Quant Title : GC Extractables
 QLast Update : Thu Sep 23 12:18:33 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds						
12) B	4,4'-DDE	5.650	6.354	54221091	297.0E6	49.812 51.491
17) MA	4,4'-DDT	6.405	7.145	11143620	64278029	10.124 12.725m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101121\
Data File : PD066156.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Oct 2021 10:30
Operator : AR\AJ
Sample : M4117-17DL 30X
Misc :
ALS Vial : 5 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
SB-5(3.0-3.5)DL


Manual Integrations
APPROVED

mohammad
10/13/2021 1:19:16 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 12 01:38:22 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD092221.M
Quant Title : GC Extractables
QLast Update : Thu Sep 23 12:18:33 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T.: 3.493 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
SB-5(3.0-3.5)DL


Manual Integrations
APPROVED

mohammad
10/13/2021 1:19:16 PM




#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T.: 4.012 min
Response: 0
Conc: N.D.




#12 4,4'-DDE

R.T.: 5.650 min
Delta R.T.: 0.002 min
Response: 54221091
Conc: 49.81 ng/ml



#12 4,4'-DDE

R.T.: 6.354 min
Delta R.T.: -0.001 min
Response: 296974774
Conc: 51.49 ng/ml




#17 4,4'-DDT

R.T.: 6.405 min
Delta R.T.: 0.002 min
Response: 11143620
Conc: 10.12 ng/ml

Instrument :
ECD_D
ClientSampleId :
SB-5(3.0-3.5)DL


Manual Integrations
APPROVED

mohammad
10/13/2021 1:19:16 PM




#17 4,4'-DDT

R.T.: 7.145 min
Delta R.T.: -0.003 min
Response: 64278029
Conc: 12.72 ng/ml m



#28 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 8.235 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 9.085 min
Response: 0
Conc: N.D.