

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Data\PD101220\  
 Data File : PD059692.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 12 Oct 2020 13:38  
 Operator : DD\AJ  
 Sample : L4337-02  
 Misc :  
 ALS Vial : 13 Sample Multiplier: 1

Instrument :  
 ECD\_D  
 ClientSampleId :  
 TP-4

Manual Integrations  
 APPROVED

Ankita  
 10/13/2020 12:26:43 PM

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Oct 12 16:13:09 2020  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Method\PD100820.M  
 Quant Title : GC Extractables  
 QLast Update : Thu Oct 08 13:21:41 2020  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 2 µl  
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1  
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

|                             | Compound        | RT#1  | RT#2  | Resp#1   | Resp#2   | ng/ml  | ng/ml   |
|-----------------------------|-----------------|-------|-------|----------|----------|--------|---------|
| -----                       |                 |       |       |          |          |        |         |
| System Monitoring Compounds |                 |       |       |          |          |        |         |
| 1)                          | SA Tetrachlo... | 3.509 | 4.025 | 29928978 | 44988435 | 20.758 | 20.694  |
| 28)                         | SA Decachlor... | 8.312 | 9.110 | 23563814 | 29966036 | 21.212 | 18.597  |
| Target Compounds            |                 |       |       |          |          |        |         |
| 17)                         | MA 4,4'-DDT     | 6.461 | 7.160 | 1034430  | 864140   | 0.679m | 0.441m# |
| -----                       |                 |       |       |          |          |        |         |

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Data\PD101220\  
Data File : PD059692.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 12 Oct 2020 13:38  
Operator : DD\AJ  
Sample : L4337-02  
Misc :  
ALS Vial : 13 Sample Multiplier: 1



Instrument :  
ECD\_D  
ClientSampled :  
TP-4


Manual Integrations  
APPROVED

Ankita  
10/13/2020 12:26:43 PM

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Oct 12 16:13:09 2020  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Method\PD100820.M  
Quant Title : GC Extractables  
QLast Update : Thu Oct 08 13:21:41 2020  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1  
Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene

R.T.: 3.509 min  
Delta R.T.: 0.000 min  
Response: 29928978  
Conc: 20.76 ng/ml

Instrument :  
ECD\_D  
ClientSampleId :  
TP-4


Manual Integrations  
APPROVED

Ankita  
10/13/2020 12:26:43 PM




#1 Tetrachloro-m-xylene

R.T.: 4.025 min  
Delta R.T.: 0.000 min  
Response: 44988435  
Conc: 20.69 ng/ml




#17 4,4'-DDT

R.T.: 6.461 min  
Delta R.T.: -0.003 min  
Response: 1034430  
Conc: 0.68 ng/ml m



#17 4,4'-DDT

R.T.: 7.160 min  
Delta R.T.: 0.002 min  
Response: 864140  
Conc: 0.44 ng/ml m




#28 Decachlorobiphenyl

R.T.: 8.312 min  
Delta R.T.: -0.004 min  
Response: 23563814  
Conc: 21.21 ng/ml

Instrument :  
ECD\_D  
ClientSampleId :  
TP-4

Manual Integrations  
APPROVED

Ankita  
10/13/2020 12:26:43 PM



#28 Decachlorobiphenyl

R.T.: 9.110 min  
Delta R.T.: 0.000 min  
Response: 29966036  
Conc: 18.60 ng/ml