

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101321\
 Data File : PD066194.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Oct 2021 22:35
 Operator : AR\AJ
 Sample : INDB401
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDB4006

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 13 02:10:28 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101321CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Oct 13 02:08:31 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.442	3.971	72113338	255.1E6	37.781	37.308
27)	SA Decachlor...	8.166	9.017	149.4E6	358.2E6	75.869	72.990
Target Compounds							
5)	MB Aldrin	4.699	5.452	111.9E6	355.4E6	38.615	37.171
6)	B beta-BHC	4.407	4.824	48644832	167.2E6	33.782	36.036
7)	B delta-BHC	4.610	5.039	103.5E6	352.0E6	38.885	36.272
8)	B Heptachlo...	5.140	5.839	98128877	299.6E6	38.445	40.446
10)	B trans-Chl...	5.364	6.073	101.2E6	361.5E6	38.870	37.007
11)	B cis-Chlor...	5.422	6.145	99857438	354.6E6	38.867	36.891
12)	B 4,4'-DDE	5.586	6.304	189.2E6	682.3E6	80.222	73.817
15)	B Endosulfa...	6.247	6.886	172.4E6	610.2E6	78.932	73.779
18)	B Endrin al...	6.415	7.010	120.2E6	440.9E6	75.939	72.679
19)	B Endosulfa...	6.627	7.232	155.6E6	561.0E6	78.595	73.654
21)	B Endrin ke...	7.119	7.704	185.1E6	587.8E6	75.710	73.428

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101321\
Data File : PD066194.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Oct 2021 22:35
Operator : AR\AJ
Sample : INDB401
Misc :
ALS Vial : 16 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 13 02:10:28 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101321CLP.M
Quant Title : GC Extractables
QLast Update : Wed Oct 13 02:08:31 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

