

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101422\
 Data File : PD072436.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Oct 2022 19:00
 Operator : AR\AJ
 Sample : N5060-16
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 15 00:28:01 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101122CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Oct 12 06:06:30 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	2.709	3.470	46990807	528.4E6	19.251	22.096
27)	SA Decachlor...	7.844	8.927	62058228	220.3E6	31.770	43.610 #
Target Compounds							
2)	A alpha-BHC	3.195	0.000	341027	0	<MDL	N.D. #
3)	MA gamma-BHC...	0.000	4.263	0 22210473		N.D.	0.854
4)	MA Heptachlor	0.000	4.827	0 41333997		N.D.	2.414
5)	MB Aldrin	4.144	0.000	1339973	0	0.407	N.D. #
6)	B beta-BHC	0.000	4.458f	0 11141067		N.D.	1.104
7)	B delta-BHC	4.046f	4.682f	764888	139.4E6	0.227	7.349 #
8)	B Heptachlo...	4.713f	0.000	148404	0	<MDL	N.D. #
10)	B trans-Chl...	4.886	5.859	120240 19770131		<MDL	1.786 #
11)	B cis-Chlor...	4.961	0.000	210938	0	<MDL	N.D. #
12)	B 4,4'-DDE	5.137	0.000	911043	0	0.316	N.D. #
13)	MA Dieldrin	5.327	0.000	231411	0	<MDL	N.D. #
14)	MA Endrin	5.519f	0.000	31413	0	<MDL	N.D. #
16)	A 4,4'-DDD	5.777f	0.000	40000	0	<MDL	N.D. #
17)	MA 4,4'-DDT	6.005	0.000	186779	0	<MDL	N.D. #
18)	B Endrin al...	6.005f	0.000	186779	0	<MDL	N.D. #
19)	B Endosulfa...	6.287	0.000	86458	0	<MDL	N.D. #
21)	B Endrin ke...	6.823	0.000	139090	0	<MDL	N.D. #
26)	Toxaphene-5	6.998	0.000	109028	0	2.722	N.D. #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101422\
Data File : PD072436.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Oct 2022 19:00
Operator : AR\AJ
Sample : N5060-16
Misc :
ALS Vial : 22 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 15 00:28:01 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101122CLP.M
Quant Title : GC Extractables
QLast Update : Wed Oct 12 06:06:30 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 2.709 min
Delta R.T.: -0.001 min
Response: 46990807
Conc: 19.25 ng/ml

Instrument :
ECD_D
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.470 min
Delta R.T.: 0.000 min
Response: 528390180
Conc: 22.10 ng/ml




#3 gamma-BHC (Lindane)

R.T.: 0.000 min
Exp R.T. : 3.546 min
Response: 0
Conc: N.D.



#3 gamma-BHC (Lindane)


R.T.: 4.263 min
Delta R.T.: -0.007 min
Response: 22210473
Conc: 0.85 ng/ml



#4 Heptachlor


R.T.: 0.000 min
Exp R.T.: 3.879 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#4 Heptachlor

R.T.: 4.827 min
Delta R.T.: -0.010 min
Response: 41333997
Conc: 2.41 ng/ml




#5 Aldrin

R.T.: 4.144 min
Delta R.T.: -0.013 min
Response: 1339973
Conc: 0.41 ng/ml



#5 Aldrin


R.T.: 0.000 min
Exp R.T.: 5.176 min
Response: 0
Conc: N.D.



#6 beta-BHC


R.T.: 0.000 min
Exp R.T.: 3.849 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#6 beta-BHC

R.T.: 4.458 min
Delta R.T.: -0.027 min
Response: 11141067
Conc: 1.10 ng/ml




#7 delta-BHC

R.T.: 4.046 min
Delta R.T.: -0.033 min
Response: 764888
Conc: 0.23 ng/ml



#7 delta-BHC


R.T.: 4.682 min
Delta R.T.: -0.048 min
Response: 139374986
Conc: 7.35 ng/ml



#10 trans-Chlordane


R.T.: 4.886 min
Delta R.T.: -0.027 min
Response: 120240
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#10 trans-Chlordane

R.T.: 5.859 min
Delta R.T.: -0.004 min
Response: 19770131
Conc: 1.79 ng/ml




#12 4,4'-DDE

R.T.: 5.137 min
Delta R.T.: -0.029 min
Response: 911043
Conc: 0.32 ng/ml



#12 4,4'-DDE


R.T.: 0.000 min
Exp R.T.: 6.111 min
Response: 0
Conc: N.D.



#26 Toxaphene-5


R.T.: 6.998 min
Delta R.T.: 0.022 min
Response: 109028
Conc: 2.72 ng/ml

Instrument :
ECD_D
ClientSampleId :




#26 Toxaphene-5

R.T.: 0.000 min
Exp R.T. : 7.499 min
Response: 0
Conc: N.D.



#27 Decachlorobiphenyl

R.T.: 7.844 min
Delta R.T.: -0.002 min
Response: 62058228
Conc: 31.77 ng/ml



#27 Decachlorobiphenyl

R.T.: 8.927 min
Delta R.T.: -0.003 min
Response: 220278289
Conc: 43.61 ng/ml