

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101520\
 Data File : PD059796.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Oct 2020 23:55
 Operator : DD\AJ
 Sample : L4308-05MSD
 Misc :
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 C0AP8MSD

Manual Integrations
 APPROVED

Ankita
 10/19/2020 10:23:45 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 16 02:32:31 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD100220CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 02 04:21:20 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.508	4.025	17324640	25737284	11.994	12.958
27)	SA Decachlor...	8.308	9.110	30990930	33333495	28.923	22.404m
Target Compounds							
3)	MA gamma-BHC...	4.236	4.701	89173512	133.5E6	45.659m	50.310
4)	MA Heptachlor	4.545	5.213	87095201	121.7E6	44.608	47.888
5)	MB Aldrin	4.796	5.520	83267874	114.9E6	44.854	47.450
10)	B trans-Chl...	5.473	6.126	5129054	32475639	3.157m	14.551 #
11)	B cis-Chlor...	5.531	6.215	6679679	16464757	4.125	7.528 #
12)	B 4,4'-DDE	5.696	6.368	2824849	15695315	1.833	7.505 #
13)	MA Dieldrin	5.831	6.523	180.6E6	224.1E6	110.089	100.517
14)	MA Endrin	6.091	6.740	135.1E6	181.1E6	114.337	116.914
16)	A 4,4'-DDD	6.215	6.862	7566831	22230294	5.788	12.841 #
17)	MA 4,4'-DDT	6.459	7.159	140.0E6	188.9E6	104.352	117.671

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101520\
Data File : PD059796.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Oct 2020 23:55
Operator : DD\AJ
Sample : L4308-05MSD
Misc :
ALS Vial : 47 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 16 02:32:31 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD100220CLP.M
Quant Title : GC Extractables
QLast Update : Fri Oct 02 04:21:20 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Manual Integrations
APPROVED

Ankita
10/19/2020 10:23:45 AM

Instrument :
ECD_D
Client Sampled :
COAP8MSD