

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101621\
 Data File : PD066289.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Oct 2021 11:17
 Operator : AR\AJ
 Sample : INDA501
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA5022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 18 23:34:27 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101621CLP.M
 Quant Title : GC Extractables
 QLast Update : Mon Oct 18 23:27:37 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.440	3.970	156.7E6	615.4E6	76.317	68.198
27)	SA Decachlor...	8.166	9.020	202.6E6	691.9E6	147.039	141.409
Target Compounds							
2)	A alpha-BHC	3.877	4.363	255.2E6	843.9E6	81.265	71.982
3)	MA gamma-BHC...	4.159	4.648	236.9E6	789.4E6	77.695	70.854
4)	MA Heptachlor	4.453	5.148	238.4E6	727.9E6	80.452	70.397
9)	A Endosulfan I	5.476	6.196	195.5E6	596.7E6	77.676	69.628
13)	MA Dieldrin	5.717	6.454	443.4E6	1251.3E6	162.817	146.441
14)	MA Endrin	5.974	6.672	368.4E6	880.7E6	155.102	156.611
16)	A 4,4'-DDD	6.099	6.798	338.6E6	991.4E6	160.538	146.802
17)	MA 4,4'-DDT	6.338	7.095	338.8E6	962.9E6	166.662	146.393
20)	A Methoxychlor	6.885	7.557	800.5E6	2187.3E6	755.018	673.305

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101621\
Data File : PD066289.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Oct 2021 11:17
Operator : AR\AJ
Sample : INDA501
Misc :
ALS Vial : 17 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 18 23:34:27 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101621CLP.M
Quant Title : GC Extractables
QLast Update : Mon Oct 18 23:27:37 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

