

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD101723CLP.M
 Title : GC Extractables
 Last Update : Tue Oct 17 06:35:07 2023
 Response Via : Initial Calibration

Calibration Files

3 =PD078914.D 1 =PD078912.D 2 =PD078913.D
 4 =PD078915.D 5 =PD078916.D

	Compound	3	1	2	4	5	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.831	1.908	1.863	1.794	1.741	1.828 E6	3.50
2) A	alpha-BHC						0.000	-1.00
3) MA	gamma-BHC (Lindane)						0.000	-1.00
4) MA	Heptachlor						0.000	-1.00
5) MB	Aldrin						0.000	-1.00
6) B	beta-BHC						0.000	-1.00
7) B	delta-BHC						0.000	-1.00
8) B	Heptachlor epoxide						0.000	-1.00
9) A	Endosulfan I						0.000	-1.00
10) B	trans-Chlordane						0.000	-1.00
11) B	cis-Chlordane						0.000	-1.00
12) B	4,4'-DDE						0.000	-1.00
13) MA	Dieldrin						0.000	-1.00
14) MA	Endrin						0.000	-1.00
15) B	Endosulfan II						0.000	-1.00
16) A	4,4'-DDD						0.000	-1.00
17) MA	4,4'-DDT						0.000	-1.00
18) B	Endrin aldehyde						0.000	-1.00
19) B	Endosulfan Sulfate						0.000	-1.00
20) A	Methoxychlor						0.000	-1.00
21) B	Endrin ketone						0.000	-1.00
22)	Toxaphene-1	7.030	7.890	7.586	6.699	6.600	7.161 E3	7.83
23)	Toxaphene-2	1.930	2.128	2.009	1.935	1.978	1.996 E4	4.03
24)	Toxaphene-3	5.668	5.868	5.714	5.778	5.840	5.774 E4	1.45
25)	Toxaphene-4	4.238	4.330	4.232	4.344	4.417	4.312 E4	1.81
26)	Toxaphene-5	3.219	3.219	3.194	3.298	3.349	3.256 E4	2.00
27) SA	Decachlorobiphenyl	2.247	2.453	2.360	2.134	2.039	2.247 E6	7.42

Signal #2 Calibration Files

3 =PD078914.D 1 =PD078912.D 2 =PD078913.D
 4 =PD078915.D 5 =PD078916.D

	Compound	3	1	2	4	5	Avg	%RSD
1) SA	Tetrachloro-m-xylene	2.576	2.659	2.618	2.564	2.489	2.581 E6	2.47
2) A	alpha-BHC						0.000	-1.00
3) MA	gamma-BHC (Lindane)						0.000	-1.00
4) MA	Heptachlor						0.000	-1.00
5) MB	Aldrin						0.000	-1.00
6) B	beta-BHC						0.000	-1.00
7) B	delta-BHC						0.000	-1.00
8) B	Heptachlor epoxide						0.000	-1.00
9) A	Endosulfan I						0.000	-1.00
10) B	trans-Chlordane						0.000	-1.00
11) B	cis-Chlordane						0.000	-1.00
12) B	4,4'-DDE						0.000	-1.00
13) MA	Dieldrin						0.000	-1.00
14) MA	Endrin						0.000	-1.00
15) B	Endosulfan II						0.000	-1.00
16) A	4,4'-DDD						0.000	-1.00
17) MA	4,4'-DDT						0.000	-1.00
18) B	Endrin aldehyde						0.000	-1.00
19) B	Endosulfan Sulfate						0.000	-1.00
20) A	Methoxychlor						0.000	-1.00
21) B	Endrin ketone						0.000	-1.00
22)	Toxaphene-1	1.552	1.598	1.553	1.534	1.484	1.544 E4	2.66
23)	Toxaphene-2	1.643	1.603	1.662	1.655	1.651	1.643 E4	1.44

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
Method File : PD101723CLP.M
Title : GC Extractables
Last Update : Tue Oct 17 06:35:07 2023
Response Via : Initial Calibration

Calibration Files

3 =PD078914.D 1 =PD078912.D 2 =PD078913.D
4 =PD078915.D 5 =PD078916.D

	Compound	3	1	2	4	5	Avg		%RSD
24)	Toxaphene-3	5.264	4.941	4.966	5.478	5.515	5.233	E4	5.21
25)	Toxaphene-4	7.262	6.937	6.987	7.372	7.352	7.182	E4	2.87
26)	Toxaphene-5	3.080	3.009	3.049	3.257	3.094	3.098	E4	3.07
27) SA	Decachlorobiphenyl	2.590	2.766	2.662	2.512	2.419	2.590	E6	5.16

(#) = Out of Range