

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101818\
 Data File : PD049845.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 18 Oct 2018 14:50
 Operator : AJ\SJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 19 07:57:01 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101818.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 19 07:43:29 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.223	4.060	12402793	549.0E6	20.688	21.392
28)	SA Decachlor...	9.006	9.167	13327790	345.6E6	21.636	21.871
Target Compounds							
2)	A alpha-BHC	4.637f	4.519f	1039717	18962006	1.230	0.509 #
6)	B beta-BHC	0.000	4.953f	0	7762590	N.D.	0.539 #
9)	A Endosulfan I	6.287f	0.000	1499046	0	2.121	N.D. #
11)	B alpha-Chl...	6.287	0.000	1499046	0	1.910	N.D. #
26)	Chlordane-4	6.287	0.000	1499046	0	20.904	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101818\
Data File : PD049845.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 18 Oct 2018 14:50
Operator : AJ\SJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 19 07:57:01 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101818.M
Quant Title : GC Extractables
QLast Update : Fri Oct 19 07:43:29 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 4.223 min
Delta R.T.: -0.002 min
Response: 12402793
Conc: 20.69 ng/ml

Instrument :
ECD_D
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.060 min
Delta R.T.: 0.000 min
Response: 548966705
Conc: 21.39 ng/ml




#2 alpha-BHC

R.T.: 4.637 min
Delta R.T.: -0.041 min
Response: 1039717
Conc: 1.23 ng/ml



#2 alpha-BHC


R.T.: 4.519 min
Delta R.T.: 0.065 min
Response: 18962006
Conc: 0.51 ng/ml



#6 beta-BHC


R.T.: 0.000 min
Exp R.T.: 5.209 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#6 beta-BHC

R.T.: 4.953 min
Delta R.T.: 0.045 min
Response: 7762590
Conc: 0.54 ng/ml




#9 Endosulfan I

R.T.: 6.287 min
Delta R.T.: -0.042 min
Response: 1499046
Conc: 2.12 ng/ml



#9 Endosulfan I


R.T.: 0.000 min
Exp R.T.: 6.306 min
Response: 0
Conc: N.D.



#11 alpha-Chlordane


R.T.: 6.287 min
Delta R.T.: 0.020 min
Response: 1499046
Conc: 1.91 ng/ml

Instrument :
ECD_D
ClientSampleId :




#11 alpha-Chlordane

R.T.: 0.000 min
Exp R.T. : 6.255 min
Response: 0
Conc: N.D.




#26 Chlordane-4

R.T.: 6.287 min
Delta R.T.: 0.021 min
Response: 1499046
Conc: 20.90 ng/ml



#26 Chlordane-4


R.T.: 0.000 min
Exp R.T. : 6.257 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 9.006 min
Delta R.T.: 0.000 min
Response: 13327790
Conc: 21.64 ng/ml

Instrument :
ECD_D
ClientSampleId :



#28 Decachlorobiphenyl

R.T.: 9.167 min
Delta R.T.: 0.000 min
Response: 345594794
Conc: 21.87 ng/ml