

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101819\
 Data File : PD055379.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Oct 2019 11:10
 Operator : SG\AJ
 Sample : INDB201
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDB201

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 18 12:57:32 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101819CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 18 12:56:12 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.290	3.963	8842775	13386497	10.233	10.377
27)	SA Decachlor...	7.967	8.999	18928705	27063942	20.731	20.571
Target Compounds							
5)	MB Aldrin	4.518	5.441	10562049	17303218	9.897	9.936
6)	B beta-BHC	4.242	4.800	5597474	9274168	10.313	10.441
7)	B delta-BHC	4.438	5.016	12224689	19030373	9.965	10.118
8)	B Heptachlo...	4.955	5.826	10136370	17318645	10.066	10.190
10)	B trans-Chl...	5.177	6.059	10376653	17791986	10.004	10.134
11)	B cis-Chlor...	5.235	6.132	10165285	17487500	10.101	10.206
12)	B 4,4'-DDE	5.403	6.287	19910745	33360145	19.374	19.703
15)	B Endosulfa...	6.053	6.861	19191464	31454938	19.963	20.131
18)	B Endrin al...	6.221	6.985	16615527	27256182	20.500	20.551
19)	B Endosulfa...	6.433	7.208	18154681	30441024	20.292	20.397
21)	B Endrin ke...	6.917	7.677	19351389	32448618	19.378	20.046


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101819\
Data File : PD055379.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Oct 2019 11:10
Operator : SG\AJ
Sample : INDB201
Misc :
ALS Vial : 12 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 18 12:57:32 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101819CLP.M
Quant Title : GC Extractables
QLast Update : Fri Oct 18 12:56:12 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.290 min
Delta R.T.: 0.000 min
Response: 8842775
Conc: 10.23 ng/ml

Instrument :
ECD_D
ClientSampleId :
INDB201




#1 Tetrachloro-m-xylene

R.T.: 3.963 min
Delta R.T.: 0.000 min
Response: 13386497
Conc: 10.38 ng/ml




#5 Aldrin

R.T.: 4.518 min
Delta R.T.: 0.000 min
Response: 10562049
Conc: 9.90 ng/ml



#5 Aldrin


R.T.: 5.441 min
Delta R.T.: -0.001 min
Response: 17303218
Conc: 9.94 ng/ml



#6 beta-BHC


R.T.: 4.242 min
Delta R.T.: 0.000 min
Response: 5597474
Conc: 10.31 ng/ml

Instrument :
ECD_D
ClientSampleId :
INDB201




#6 beta-BHC

R.T.: 4.800 min
Delta R.T.: 0.000 min
Response: 9274168
Conc: 10.44 ng/ml




#7 delta-BHC

R.T.: 4.438 min
Delta R.T.: 0.000 min
Response: 12224689
Conc: 9.96 ng/ml



#7 delta-BHC


R.T.: 5.016 min
Delta R.T.: 0.000 min
Response: 19030373
Conc: 10.12 ng/ml



#8 Heptachlor epoxide


R.T.: 4.955 min
Delta R.T.: 0.000 min
Response: 10136370
Conc: 10.07 ng/ml

Instrument :
ECD_D
ClientSampleId :
INDB201




#8 Heptachlor epoxide

R.T.: 5.826 min
Delta R.T.: 0.000 min
Response: 17318645
Conc: 10.19 ng/ml




#10 trans-Chlordane

R.T.: 5.177 min
Delta R.T.: 0.000 min
Response: 10376653
Conc: 10.00 ng/ml



#10 trans-Chlordane


R.T.: 6.059 min
Delta R.T.: 0.000 min
Response: 17791986
Conc: 10.13 ng/ml



#11 cis-Chlordane


R.T.: 5.235 min
Delta R.T.: 0.000 min
Response: 10165285
Conc: 10.10 ng/ml

Instrument :
ECD_D
ClientSampleId :
INDB201




#11 cis-Chlordane

R.T.: 6.132 min
Delta R.T.: 0.000 min
Response: 17487500
Conc: 10.21 ng/ml




#12 4,4'-DDE

R.T.: 5.403 min
Delta R.T.: 0.000 min
Response: 19910745
Conc: 19.37 ng/ml



#12 4,4'-DDE


R.T.: 6.287 min
Delta R.T.: -0.001 min
Response: 33360145
Conc: 19.70 ng/ml



#15 Endosulfan II


R.T.: 6.053 min
Delta R.T.: 0.000 min
Response: 19191464
Conc: 19.96 ng/ml

Instrument :
ECD_D
ClientSampleId :
INDB201




#15 Endosulfan II

R.T.: 6.861 min
Delta R.T.: 0.000 min
Response: 31454938
Conc: 20.13 ng/ml




#18 Endrin aldehyde

R.T.: 6.221 min
Delta R.T.: 0.000 min
Response: 16615527
Conc: 20.50 ng/ml



#18 Endrin aldehyde


R.T.: 6.985 min
Delta R.T.: 0.000 min
Response: 27256182
Conc: 20.55 ng/ml



#19 Endosulfan Sulfate


R.T.: 6.433 min
Delta R.T.: 0.001 min
Response: 18154681
Conc: 20.29 ng/ml

Instrument :
ECD_D
ClientSampleId :
INDB201




#19 Endosulfan Sulfate

R.T.: 7.208 min
Delta R.T.: 0.000 min
Response: 30441024
Conc: 20.40 ng/ml




#21 Endrin ketone

R.T.: 6.917 min
Delta R.T.: 0.000 min
Response: 19351389
Conc: 19.38 ng/ml



#21 Endrin ketone


R.T.: 7.677 min
Delta R.T.: 0.000 min
Response: 32448618
Conc: 20.05 ng/ml



#27 Decachlorobiphenyl

R.T.: 7.967 min
Delta R.T.: 0.000 min
Response: 18928705
Conc: 20.73 ng/ml

Instrument :
ECD_D
ClientSampleId :
INDB201



#27 Decachlorobiphenyl

R.T.: 8.999 min
Delta R.T.: -0.001 min
Response: 27063942
Conc: 20.57 ng/ml