

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101823\
 Data File : PD078959.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Oct 2023 18:16
 Operator : AR\AJ
 Sample : 04885-05MS 10X
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 COM28MS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 17 22:14:49 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101623CLP-TCLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Oct 17 05:35:36 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.568	2.806	32976856	40990746	21.428	19.170
2)	SA Decachlor...	9.116	7.957	43824501	51351957	24.001	22.685
Target Compounds							
2)	A alpha-BHC	4.026	3.310	141.1E6	179.1E6	53.006	53.662
3)	MA gamma-BHC...	4.359	3.640	126.0E6	159.5E6	49.670	51.205
4)	MA Heptachlor	4.953	3.983	127.3E6	149.3E6	47.042	50.041
5)	MB Aldrin	5.296	4.264	120.5E6	150.3E6	48.715	49.215
6)	B beta-BHC	4.552	3.934	54134747	68191489	47.984	47.715
7)	B delta-BHC	4.802	4.165	95673462	117.7E6	35.401	36.049
8)	B Heptachlo...	5.722	4.766	117.7E6	150.2E6	51.221	52.640
9)	A Endosulfan I	6.109	5.137	111.4E6	138.5E6	52.768	54.647
10)	B trans-Chl...	5.977	5.017	113.7E6	144.0E6	51.312	52.159
11)	B cis-Chlor...	6.058	5.081	114.2E6	145.8E6	50.679	51.213
12)	B 4,4'-DDE	6.228	5.271	220.9E6	282.1E6	106.445	107.888
13)	MA Dieldrin	6.385	5.402	241.3E6	310.8E6	106.068	108.516
14)	MA Endrin	6.615	5.677	203.3E6	273.2E6	107.990	109.604
15)	B Endosulfa...	6.832	5.969	204.2E6	258.2E6	105.027	109.294
16)	A 4,4'-DDD	6.746	5.824	179.3E6	232.5E6	106.263	106.161
17)	MA 4,4'-DDT	7.061	6.076	192.8E6	239.1E6	106.866	106.605
18)	B Endrin al...	6.963	6.148	146.2E6	184.8E6	97.534	101.397
19)	B Endosulfa...	7.197	6.371	186.5E6	227.2E6	98.383	100.516
20)	A Methoxychlor	7.535	6.651	495.0E6	601.0E6	519.193	512.735
21)	B Endrin ke...	7.683	6.877	214.0E6	278.3E6	103.940	106.940

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101823\
Data File : PD078959.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Oct 2023 18:16
Operator : AR\AJ
Sample : 04885-05MS 10X
Misc :
ALS Vial : 16 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 22:14:49 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101623CLP-TCLP.M
Quant Title : GC Extractables
QLast Update : Tue Oct 17 05:35:36 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

