

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD101918CLP.M
 Title : GC Extractables
 Last Update : Fri Oct 19 03:46:25 2018
 Response Via : Initial Calibration

Calibration Files

3 =PD049871.D 1 =PD049869.D 2 =PD049870.D
 4 =PD049872.D 5 =PD049873.D

Compound			3	1	2	4	5	Avg		%RSD
1) SA	Tetrachloro-m-xylene		6.631	6.849	6.718	6.497	6.464	6.632	E5	2.39
2) A	alpha-BHC							0.000		-1.00
3) MA	gamma-BHC (Lindane)							0.000		-1.00
4) MA	Heptachlor							0.000		-1.00
5) MB	Aldrin							0.000		-1.00
6) B	beta-BHC							0.000		-1.00
7) B	delta-BHC							0.000		-1.00
8) B	Heptachlor epoxide							0.000		-1.00
9) A	Endosulfan I							0.000		-1.00
10) B	trans-Chlordane							0.000		-1.00
11) B	cis-Chlordane							0.000		-1.00
12) B	4,4'-DDE							0.000		-1.00
13) MA	Dieldrin							0.000		-1.00
14) MA	Endrin							0.000		-1.00
15) B	Endosulfan II							0.000		-1.00
16) A	4,4'-DDD							0.000		-1.00
17) MA	4,4'-DDT							0.000		-1.00
18) B	Endrin aldehyde							0.000		-1.00
19) B	Endosulfan Sulfate							0.000		-1.00
20) A	Methoxychlor							0.000		-1.00
21) B	Endrin ketone							0.000		-1.00
22)	Toxaphene-1		5.885	6.112	6.213	5.923	5.894	6.005	E3	2.48
23)	Toxaphene-2		6.405	6.462	6.475	6.299	6.572	6.442	E3	1.56
24)	Toxaphene-3		1.325	1.331	1.313	1.360	1.370	1.340	E4	1.82
25)	Toxaphene-4		1.454	1.448	1.444	1.516	1.519	1.476	E4	2.55
26)	Toxaphene-5		1.318	1.282	1.282	1.296	1.326	1.301	E4	1.56
27) SA	Decachlorobiphenyl		6.727	7.389	7.082	6.398	6.125	6.744	E5	7.54

Signal #2 Calibration Files

3 =PD049871.D 1 =PD049869.D 2 =PD049870.D
 4 =PD049872.D 5 =PD049873.D

Compound			3	1	2	4	5	Avg		%RSD
1) SA	Tetrachloro-m-xylene		3.078	3.323	3.218	2.863	2.697	3.036	E7	8.44
2) A	alpha-BHC							0.000		-1.00
3) MA	gamma-BHC (Lindane)							0.000		-1.00
4) MA	Heptachlor							0.000		-1.00
5) MB	Aldrin							0.000		-1.00
6) B	beta-BHC							0.000		-1.00
7) B	delta-BHC							0.000		-1.00
8) B	Heptachlor epoxide							0.000		-1.00
9) A	Endosulfan I							0.000		-1.00
10) B	trans-Chlordane							0.000		-1.00
11) B	cis-Chlordane							0.000		-1.00
12) B	4,4'-DDE							0.000		-1.00
13) MA	Dieldrin							0.000		-1.00
14) MA	Endrin							0.000		-1.00
15) B	Endosulfan II							0.000		-1.00
16) A	4,4'-DDD							0.000		-1.00
17) MA	4,4'-DDT							0.000		-1.00
18) B	Endrin aldehyde							0.000		-1.00
19) B	Endosulfan Sulfate							0.000		-1.00
20) A	Methoxychlor							0.000		-1.00
21) B	Endrin ketone							0.000		-1.00
22)	Toxaphene-1		1.249	1.447	1.342	1.155	1.074	1.253	E5	11.78
23)	Toxaphene-2		1.478	1.602	1.521	1.351	1.266	1.443	E5	9.31

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD101918CLP.M
 Title : GC Extractables
 Last Update : Fri Oct 19 03:46:25 2018
 Response Via : Initial Calibration

Calibration Files

3 =PD049871.D 1 =PD049869.D 2 =PD049870.D
 4 =PD049872.D 5 =PD049873.D

	Compound	3	1	2	4	5	Avg		%RSD
24)	Toxaphene-3	1.843	1.961	1.875	1.739	1.693	1.822	E5	5.89
25)	Toxaphene-4	1.116	1.124	1.148	1.115	1.104	1.121	E5	1.45
26)	Toxaphene-5	2.106	2.279	2.176	2.116	2.054	2.146	E5	4.01
27) SA	Decachlorobiphenyl	1.786	2.074	1.936	1.635	1.493	1.785	E7	12.95

(#) = Out of Range