

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD101918CLP.M
 Title : GC Extractables
 Last Update : Fri Oct 19 04:21:11 2018
 Response Via : Initial Calibration

Calibration Files

3 =PD049878.D 1 =PD049874.D 2 =PD049876.D
 4 =PD049880.D 5 =PD049882.D

Compound		3	1	2	4	5	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	6.543	6.782	6.708	6.598	6.441	6.614	E5	2.03
2) A	alpha-BHC	8.772	8.731	8.778	9.240	9.377	8.980	E5	3.39
3) MA	gamma-BHC (Lindane)	8.900	9.779	9.286	9.095	9.034	9.219	E5	3.71
4) MA	Heptachlor	8.616	8.569	8.735	8.832	8.849	8.720	E5	1.44
5) MB	Aldrin						0.000		-1.00
6) B	beta-BHC						0.000		-1.00
7) B	delta-BHC						0.000		-1.00
8) B	Heptachlor epoxide						0.000		-1.00
9) A	Endosulfan I	7.680	7.769	7.864	7.766	7.650	7.746	E5	1.09
10) B	trans-Chlordane						0.000		-1.00
11) B	cis-Chlordane						0.000		-1.00
12) B	4,4'-DDE						0.000		-1.00
13) MA	Dieldrin	8.574	8.199	8.462	8.910	8.983	8.626	E5	3.76
14) MA	Endrin	6.291	6.267	6.299	6.412	6.134	6.280	E5	1.58
15) B	Endosulfan II						0.000		-1.00
16) A	4,4'-DDD	6.416	5.809	6.246	6.699	6.734	6.381	E5	5.93
17) MA	4,4'-DDT	5.951	5.395	5.766	6.318	6.435	5.973	E5	7.05
18) B	Endrin aldehyde						0.000		-1.00
19) B	Endosulfan Sulfate						0.000		-1.00
20) A	Methoxychlor	3.493	3.358	3.486	3.587	3.495	3.484	E5	2.34
21) B	Endrin ketone						0.000		-1.00
22)	Toxaphene-1						0.000		-1.00
23)	Toxaphene-2						0.000		-1.00
24)	Toxaphene-3						0.000		-1.00
25)	Toxaphene-4						0.000		-1.00
26)	Toxaphene-5						0.000		-1.00
27) SA	Decachlorobiphenyl	6.628	7.186	7.065	6.447	6.082	6.681	E5	6.77

Signal #2 Calibration Files

3 =PD049878.D 1 =PD049874.D 2 =PD049876.D
 4 =PD049880.D 5 =PD049882.D

Compound		3	1	2	4	5	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	3.020	3.297	3.230	2.889	2.688	3.025	E7	8.24
2) A	alpha-BHC	4.459	4.877	4.715	4.325	4.062	4.488	E7	7.15
3) MA	gamma-BHC (Lindane)	4.024	4.347	4.256	3.880	3.640	4.030	E7	7.09
4) MA	Heptachlor	3.938	4.348	4.201	3.732	3.444	3.933	E7	9.20
5) MB	Aldrin						0.000		-1.00
6) B	beta-BHC						0.000		-1.00
7) B	delta-BHC						0.000		-1.00
8) B	Heptachlor epoxide						0.000		-1.00
9) A	Endosulfan I	2.895	3.167	3.072	2.741	2.529	2.881	E7	8.88
10) B	trans-Chlordane						0.000		-1.00
11) B	cis-Chlordane						0.000		-1.00
12) B	4,4'-DDE						0.000		-1.00
13) MA	Dieldrin	2.882	3.080	3.023	2.789	2.607	2.876	E7	6.57
14) MA	Endrin	2.052	2.150	2.132	1.998	1.784	2.023	E7	7.28
15) B	Endosulfan II			3.216			3.216	E4	0.00
16) A	4,4'-DDD	2.118	2.230	2.208	2.057	1.921	2.107	E7	5.93
17) MA	4,4'-DDT	1.571	1.533	1.574	1.599	1.565	1.568	E7	1.51
18) B	Endrin aldehyde						0.000		-1.00
19) B	Endosulfan Sulfate						0.000		-1.00
20) A	Methoxychlor	7.968	8.086	8.108	7.975	7.490	7.925	E6	3.17
21) B	Endrin ketone						0.000		-1.00
22)	Toxaphene-1						0.000		-1.00
23)	Toxaphene-2						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD101918CLP.M
 Title : GC Extractables
 Last Update : Fri Oct 19 04:21:11 2018
 Response Via : Initial Calibration

Calibration Files

3 =PD049878.D 1 =PD049874.D 2 =PD049876.D
 4 =PD049880.D 5 =PD049882.D

	Compound	3	1	2	4	5	Avg	%RSD
24)	Toxaphene-3						0.000	-1.00
25)	Toxaphene-4						0.000	-1.00
26)	Toxaphene-5	2.834					2.834 E3	0.00
27) SA	Decachlorobiphenyl	1.773	2.029	1.933	1.660	1.489	1.777 E7	12.08

(#) = Out of Range