

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD101918CLP.M
 Title : GC Extractables
 Last Update : Fri Oct 19 05:01:26 2018
 Response Via : Initial Calibration

Calibration Files

3 =PD049879.D 1 =PD049875.D 2 =PD049877.D
 4 =PD049881.D 5 =PD049883.D

Compound			3	1	2	4	5	Avg		%RSD
1) SA	Tetrachloro-m-xylene		6.131	6.311	6.382	6.169	6.024	6.203	E5	2.31
2) A	alpha-BHC							0.000		-1.00
3) MA	gamma-BHC (Lindane)							0.000		-1.00
4) MA	Heptachlor							0.000		-1.00
5) MB	Aldrin		7.941	7.804	8.065	8.195	8.204	8.042	E5	2.13
6) B	beta-BHC		4.055	4.283	4.226	3.968	3.766	4.059	E5	5.11
7) B	delta-BHC		8.262	7.985	8.254	8.635	8.735	8.374	E5	3.66
8) B	Heptachlor epoxide		7.689	7.788	7.909	7.775	7.659	7.764	E5	1.26
9) A	Endosulfan I							0.000		-1.00
10) B	trans-Chlordane		8.104	9.193	8.638	8.094	7.912	8.389	E5	6.26
11) B	cis-Chlordane		7.919	9.042	8.465	7.854	7.620	8.180	E5	7.00
12) B	4,4'-DDE		7.363	7.027	7.231	7.708	7.794	7.425	E5	4.35
13) MA	Dieldrin							0.000		-1.00
14) MA	Endrin							0.000		-1.00
15) B	Endosulfan II		7.241	7.378	7.383	7.323	7.223	7.310	E5	1.03
16) A	4,4'-DDD							0.000		-1.00
17) MA	4,4'-DDT							0.000		-1.00
18) B	Endrin aldehyde		5.439	5.603	5.602	5.448	5.315	5.482	E5	2.23
19) B	Endosulfan Sulfate		6.779	6.980	6.947	6.800	6.668	6.835	E5	1.88
20) A	Methoxychlor							0.000		-1.00
21) B	Endrin ketone		7.055	7.012	7.107	7.130	6.995	7.060	E5	0.83
22)	Toxaphene-1							0.000		-1.00
23)	Toxaphene-2							0.000		-1.00
24)	Toxaphene-3							0.000		-1.00
25)	Toxaphene-4							0.000		-1.00
26)	Toxaphene-5							0.000		-1.00
27) SA	Decachlorobiphenyl		6.296	6.795	6.625	6.128	5.762	6.321	E5	6.46

Signal #2 Calibration Files

3 =PD049879.D 1 =PD049875.D 2 =PD049877.D
 4 =PD049881.D 5 =PD049883.D

	Compound	3	1	2	4	5	Avg		%RSD
1)	SA Tetrachloro-m-xylene	2.861	3.117	3.029	2.729	2.540	2.855	E7	8.10
2)	A alpha-BHC						0.000		-1.00
3)	MA gamma-BHC (Lindane)						0.000		-1.00
4)	MA Heptachlor						0.000		-1.00
5)	MB Aldrin	3.465	3.743	3.661	3.305	3.064	3.447	E7	7.95
6)	B beta-BHC	1.846	2.154	2.001	1.706	1.549	1.851	E7	12.86
7)	B delta-BHC	3.832	4.141	4.019	3.685	3.464	3.828	E7	7.00
8)	B Heptachlor epoxide	3.111	3.463	3.318	2.938	2.702	3.106	E7	9.70
9)	A Endosulfan I						0.000		-1.00
10)	B trans-Chlordane	3.055	3.354	3.239	2.904	2.707	3.052	E7	8.46
11)	B cis-Chlordane	2.898	3.175	3.080	2.747	2.548	2.889	E7	8.73
12)	B 4,4'-DDE	2.657	2.833	2.773	2.572	2.415	2.650	E7	6.26
13)	MA Dieldrin						0.000		-1.00
14)	MA Endrin						0.000		-1.00
15)	B Endosulfan II	2.220	2.409	2.343	2.121	1.971	2.213	E7	7.89
16)	A 4,4'-DDD						0.000		-1.00
17)	MA 4,4'-DDT						0.000		-1.00
18)	B Endrin aldehyde	1.714	1.878	1.819	1.634	1.513	1.712	E7	8.51
19)	B Endosulfan Sulfate	2.020	2.217	2.141	1.929	1.789	2.019	E7	8.41
20)	A Methoxychlor						0.000		-1.00
21)	B Endrin ketone	1.834	1.954	1.913	1.768	1.662	1.826	E7	6.39
22)	Toxaphene-1						0.000		-1.00
23)	Toxaphene-2						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
Method File : PD101918CLP.M
Title : GC Extractables
Last Update : Fri Oct 19 05:01:26 2018
Response Via : Initial Calibration

Calibration Files

3 =PD049879.D 1 =PD049875.D 2 =PD049877.D
4 =PD049881.D 5 =PD049883.D

	Compound	3	1	2	4	5	Avg	%RSD
24)	Toxaphene-3						0.000	-1.00
25)	Toxaphene-4						0.000	-1.00
26)	Toxaphene-5						0.000	-1.00
27) SA	Decachlorobiphenyl	1.687	1.901	1.814	1.563	1.414	1.676 E7	11.60

(#) = Out of Range