

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : DTX101920.M
 Title : GC Extractables
 Last Update : Tue Oct 20 07:47:14 2020
 Response Via : Initial Calibration

Calibration Files

500 =PD059825.D 1000 =PD059823.D 750 =PD059824.D
 250 =PD059826.D 100 =PD059827.D

	Compound	500	1000	750	250	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	1.480	1.430	1.462	1.488	1.518	1.476	E6	2.20
2)	Toxaphene-1	1.849	1.698	1.768	2.050	2.267	1.927	E4	12.02
3)	Toxaphene-2	3.053	2.923	3.015	3.370	4.073	3.287	E4	14.32
4)	Toxaphene-3	3.008	3.068	3.107	3.051	2.992	3.045	E4	1.51
5)	Toxaphene-4	1.128	1.141	1.131	1.169	1.164	1.147	E4	1.63
6)	Toxaphene-5	1.293	1.222	1.253	1.325	1.451	1.309	E4	6.78
7) SA	Decachlorobiphenyl	1.558	1.481	1.524	1.614	1.670	1.569	E6	4.74

Signal #2 Calibration Files

500 =PD059825.D 1000 =PD059823.D 750 =PD059824.D
 250 =PD059826.D 100 =PD059827.D

	Compound	500	1000	750	250	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	2.454	2.398	2.458	2.442	2.444	2.439	E6	0.98
2)	Toxaphene-1	2.606	2.451	2.599	2.690	2.729	2.615	E4	4.10
3)	Toxaphene-2	2.608	2.482	2.561	2.691	2.788	2.626	E4	4.50
4)	Toxaphene-3	6.096	5.820	6.058	6.329	6.568	6.174	E4	4.61
5)	Toxaphene-4	3.827	3.678	3.819	3.968	4.006	3.860	E4	3.40
6)	Toxaphene-5	2.704	2.612	2.905	2.764	2.821	2.761	E4	4.04
7) SA	Decachlorobiphenyl	1.832	1.734	1.789	1.886	1.907	1.830	E6	3.87

(#) = Out of Range