

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102022\
 Data File : PD072531.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Oct 2022 17:34
 Operator : AR\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PEM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 05:03:04 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102022.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 21 04:17:06 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	2.706	3.466	48706929	692.7E6	19.548	20.830
28)	SA Decachlor...	7.842	8.925	44983751	180.6E6	19.986	20.077
Target Compounds							
2)	A alpha-BHC	3.211	3.932	36084118	491.9E6	9.163	10.743
3)	MA gamma-BHC...	3.541	4.265	28787675	413.3E6	8.824	11.232 #
6)	B beta-BHC	3.844	4.481	16010106	160.2E6	9.947	11.272
14)	MA Endrin	5.572	6.497	126.3E6	592.8E6	46.507	44.385
16)	A 4,4'-DDD	5.720	6.633	511867	9726865	0.209	0.789 #
17)	MA 4,4'-DDT	5.968	6.939	245.1E6	1094.2E6	104.660	91.788
18)	B Endrin al...	6.054	6.859	3069748	12305230	1.561	1.275
20)	A Methoxychlor	6.552	7.419	327.3E6	1308.4E6	243.290	220.969
21)	B Endrin ke...	6.786	7.581	9250601	30550279	3.097	2.448

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102022\
Data File : PD072531.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2022 17:34
Operator : AR\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PEM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 05:03:04 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102022.M
Quant Title : GC Extractables
QLast Update : Fri Oct 21 04:17:06 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

