

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102022\
 Data File : PD072553.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Oct 2022 22:27
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 21 04:11:20 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102022.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 21 02:43:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	2.705	3.466	121.2E6	1551.8E6	48.648	46.676
28)	SA Decachlor...	7.840	8.923	110.9E6	447.0E6	49.285	49.685
Target Compounds							
2)	A alpha-BHC	3.210	3.931	200.6E6	2185.0E6	50.950	47.719
3)	MA gamma-BHC...	3.541	4.265	174.3E6	1731.3E6	53.413	47.251
4)	MA Heptachlor	3.873	4.831	180.6E6	1212.9E6	50.489	46.414
5)	MB Aldrin	4.152	5.170	174.9E6	996.9E6	50.656	47.274
6)	B beta-BHC	3.843	4.480	72884281	663.8E6	51.982	46.604
7)	B delta-BHC	4.073	4.724	181.7E6	1357.2E6	51.074	47.856
8)	B Heptachlo...	4.658	5.603	154.4E6	818.3E6	50.278	46.836
9)	A Endosulfan I	5.027	5.987	144.8E6	738.8E6	50.261	48.014
10)	B gamma-Chl...	4.907	5.857	160.2E6	788.0E6	50.526	47.827
11)	B alpha-Chl...	4.970	5.932	155.7E6	773.4E6	50.172	48.197
12)	B 4,4'-DDE	5.159	6.105	151.5E6	740.7E6	50.727	48.399
13)	MA Dieldrin	5.293	6.264	154.8E6	767.2E6	50.652	48.548
14)	MA Endrin	5.570	6.496	132.5E6	621.9E6	48.802	46.893
15)	B Endosulfa...	5.869	6.726	123.2E6	607.1E6	50.387	49.392
16)	A 4,4'-DDD	5.717	6.631	125.9E6	608.0E6	51.302	49.346
17)	MA 4,4'-DDT	5.967	6.938	118.4E6	590.7E6	50.548	49.552
18)	B Endrin al...	6.051	6.857	97687982	477.3E6	49.698	49.451
19)	B Endosulfa...	6.274	7.090	123.0E6	572.0E6	48.919	50.173
20)	A Methoxychlor	6.550	7.418	65701999	290.6E6	48.562	49.070
21)	B Endrin ke...	6.785	7.580	149.1E6	607.9E6	49.897	48.717
22)	Mirex	6.949	8.028	104.7E6	413.0E6	49.242	49.679

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102022\
Data File : PD072553.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Oct 2022 22:27
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 21 04:11:20 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102022.M
Quant Title : GC Extractables
QLast Update : Fri Oct 21 02:43:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

