

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102023\
 Data File : PD078989.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Oct 2023 18:21
 Operator : AR\AJ
 Sample : PSTDICC100
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC100

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 19 19:08:26 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102023.M
 Quant Title : GC Extractables
 QLast Update : Thu Oct 19 19:02:36 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.569	2.806	159.5E6	228.1E6	100.268	100.290
28)	SA Decachlor...	9.116	7.957	181.3E6	228.3E6	97.443	98.763
Target Compounds							
2)	A alpha-BHC	4.027	3.310	282.2E6	360.9E6	102.707	101.829
3)	MA gamma-BHC...	4.360	3.640	269.4E6	349.3E6	101.810	101.319
4)	MA Heptachlor	4.953	3.983	269.3E6	310.2E6	100.873	101.077
5)	MB Aldrin	5.297	4.265	259.1E6	330.6E6	100.935	100.834
6)	B beta-BHC	4.552	3.934	101.4E6	135.0E6	98.720	99.113
7)	B delta-BHC	4.802	4.166	281.2E6	346.1E6	101.647	101.541
8)	B Heptachlo...	5.723	4.767	228.8E6	294.7E6	100.080	100.246
9)	A Endosulfan I	6.110	5.137	210.1E6	264.1E6	99.892	100.492
10)	B gamma-Chl...	5.978	5.017	226.5E6	289.0E6	100.728	100.998
11)	B alpha-Chl...	6.058	5.082	222.8E6	285.9E6	100.201	100.581
12)	B 4,4'-DDE	6.229	5.271	214.4E6	273.4E6	101.466	101.584
13)	MA Dieldrin	6.385	5.402	232.9E6	302.3E6	101.022	101.165
14)	MA Endrin	6.615	5.677	193.5E6	257.1E6	100.730	100.750
15)	B Endosulfa...	6.832	5.969	195.8E6	246.8E6	99.899	100.306
16)	A 4,4'-DDD	6.746	5.824	176.3E6	231.0E6	101.159	101.125
17)	MA 4,4'-DDT	7.061	6.076	188.6E6	234.3E6	100.567	101.187
18)	B Endrin al...	6.962	6.149	146.3E6	180.3E6	98.507	99.201
19)	B Endosulfa...	7.197	6.371	189.3E6	227.9E6	99.695	99.899
20)	A Methoxychlor	7.535	6.652	101.2E6	128.1E6	98.152	98.410
21)	B Endrin ke...	7.684	6.878	208.6E6	273.8E6	99.612	99.745
22)	Mirex	8.160	7.064	150.4E6	199.5E6	97.271	98.410

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102023\
Data File : PD078989.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Oct 2023 18:21
Operator : AR\AJ
Sample : PSTDICC100
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC100

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 19 19:08:26 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102023.M
Quant Title : GC Extractables
QLast Update : Thu Oct 19 19:02:36 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

