

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102023\
 Data File : PD079009.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Oct 2023 02:49
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 10/20/2023
 Supervised By :Ankita Jodhani 10/20/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 20 04:47:27 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102023.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 20 04:44:01 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.570	2.807	82357503	120.0E6	51.009	51.838
28)	SA Decachlor...	9.117	7.956	98905455	126.7E6	49.304	51.960
Target Compounds							
2)	A alpha-BHC	4.027	3.310	139.8E6	186.4E6	52.531	53.746
3)	MA gamma-BHC...	4.361	3.640	135.0E6	182.5E6	51.952	53.744
4)	MA Heptachlor	4.954	3.984	134.3E6	163.9E6	50.210	53.676
5)	MB Aldrin	5.297	4.265	131.8E6	174.9E6	51.141	53.466
6)	B beta-BHC	4.553	3.935	53925702	73811823	50.323	52.382
7)	B delta-BHC	4.803	4.167	134.2E6	180.9E6	50.139	54.024
8)	B Heptachlo...	5.724	4.767	119.7E6	157.9E6	51.055	53.253
9)	A Endosulfan I	6.110	5.137	109.8E6	137.0E6	50.650	51.749
10)	B gamma-Chl...	5.979	5.017	116.9E6	150.8E6	50.990	52.570
11)	B alpha-Chl...	6.059	5.082	115.6E6	148.7E6	50.705	51.807
12)	B 4,4'-DDE	6.230	5.271	108.5E6	144.1E6	51.359	54.125
13)	MA Dieldrin	6.385	5.402	119.0E6	159.7E6	51.114	53.777
14)	MA Endrin	6.616	5.677	98141921	136.2E6	50.391	53.209
15)	B Endosulfa...	6.832	5.970	102.2E6	133.0E6	50.097	53.258
16)	A 4,4'-DDD	6.746	5.824	90008416	122.8E6	51.201	53.759
17)	MA 4,4'-DDT	7.061	6.076	94833521	123.2E6	50.079	53.826
18)	B Endrin al...	6.962	6.149	78691244	100.4E6	50.956m	53.987
19)	B Endosulfa...	7.197	6.372	99550057	123.9E6	50.072	53.049
20)	A Methoxychlor	7.535	6.651	54494966	70852190	49.732	52.188
21)	B Endrin ke...	7.684	6.878	110.9E6	149.7E6	50.869	53.538
22)	Mirex	8.161	7.064	83028051	111.6E6	49.052	52.251

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102023\
 Data File : PD079009.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Oct 2023 02:49
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

PSTDCCC050

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 10/20/2023

Supervised By :Ankita Jodhani 10/20/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 20 04:47:27 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102023.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 20 04:44:01 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

