

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102124\
 Data File : PD086153.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2024 22:32
 Operator : AR\AJ
 Sample : PSTDICV050
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 ICVPD102124

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 10/22/2024
 Supervised By :Ankita Jodhani 10/22/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 22 02:07:22 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102124.M
 Quant Title : GC Extractables
 QLast Update : Tue Oct 22 02:04:00 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.381	2.822	84749206	514.2E6	50.867	51.613
28)	SA Decachlor...	8.815	7.977	99607563	532.6E6	49.555	51.657
Target Compounds							
2)	A alpha-BHC	3.824	3.326	189.7E6	851.3E6	51.331	51.921
3)	MA gamma-BHC...	4.151	3.658	177.3E6	802.0E6	50.274	51.936
4)	MA Heptachlor	4.743	4.006	144.9E6	751.4E6	50.608	52.086
5)	MB Aldrin	5.081	4.290	170.5E6	786.1E6	51.099	51.948
6)	B beta-BHC	4.340	3.952	62681547	311.7E6	51.576	51.375
7)	B delta-BHC	4.584	4.186	178.1E6	811.8E6	51.162	51.621
8)	B Heptachlo...	5.500	4.791	151.6E6	681.4E6	52.048	51.522
9)	A Endosulfan I	5.879	5.163	137.8E6	635.5E6	50.864	51.615
10)	B gamma-Chl...	5.753	5.043	144.8E6	718.3E6	50.835	51.914
11)	B alpha-Chl...	5.833	5.107	144.5E6	693.8E6	50.918	51.892
12)	B 4,4'-DDE	6.007	5.291	128.2E6	692.2E6	50.939	51.758
13)	MA Dieldrin	6.150	5.428	146.0E6	716.7E6	50.788	51.636
14)	MA Endrin	6.373	5.703	118.9E6	659.0E6	50.659m	51.770
15)	B Endosulfa...	6.587	5.993	130.9E6	602.8E6	55.683	51.246m
16)	A 4,4'-DDD	6.514	5.844	102.2E6	581.9E6	51.396	52.062
17)	MA 4,4'-DDT	6.827	6.097	87366583	605.2E6	52.409	52.033
18)	B Endrin al...	6.716	6.172	90542131	497.8E6	50.802	53.064
19)	B Endosulfa...	6.950	6.395	100.7E6	615.4E6	51.018	51.884
20)	A Methoxychlor	7.302	6.666	43482510	296.4E6	51.229	52.378
21)	B Endrin ke...	7.425	6.903	117.1E6	674.7E6	51.768	52.441
22)	Mirex	7.903	7.093	84248100	537.3E6	53.800	53.828

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102124\
Data File : PD086153.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2024 22:32
Operator : AR\AJ
Sample : PSTDICV050
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

ICVPD102124

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 10/22/2024

Supervised By :Ankita Jodhani 10/22/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 22 02:07:22 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102124.M
Quant Title : GC Extractables
QLast Update : Tue Oct 22 02:04:00 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

