

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102220\
 Data File : PD059897.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Oct 2020 20:19
 Operator : DD\AJ
 Sample : INDA501
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA501

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 22 02:21:39 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102220CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu Oct 22 02:19:57 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.508	4.024	111.2E6	200.4E6	75.663	77.945
27)	SA Decachlor...	8.305	9.109	268.4E6	311.1E6	145.643	147.685
Target Compounds							
2)	A alpha-BHC	3.949	4.415	180.1E6	308.2E6	80.255	81.244
3)	MA gamma-BHC...	4.236	4.701	168.5E6	281.2E6	77.516	81.190
4)	MA Heptachlor	4.543	5.212	173.5E6	269.6E6	78.547	79.498
9)	A Endosulfan I	5.584	6.263	146.9E6	215.5E6	77.747	77.616
13)	MA Dieldrin	5.829	6.521	337.6E6	483.4E6	159.194	158.988
14)	MA Endrin	6.088	6.739	268.8E6	380.6E6	156.156	158.361
16)	A 4,4'-DDD	6.213	6.858	268.1E6	382.9E6	159.186	159.314
17)	MA 4,4'-DDT	6.456	7.157	292.0E6	387.4E6	160.623	162.952
20)	A Methoxychlor	7.005	7.614	671.8E6	894.6E6	722.896	728.090

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102220\
Data File : PD059897.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Oct 2020 20:19
Operator : DD\AJ
Sample : INDA501
Misc :
ALS Vial : 17 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 22 02:21:39 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102220CLP.M
Quant Title : GC Extractables
QLast Update : Thu Oct 22 02:19:57 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

