

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102224\
 Data File : PD086182.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Oct 2024 18:49
 Operator : AR\AJ
 Sample : INDA304
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA3113

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 23 03:37:29 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102224CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Oct 22 16:12:51 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.379	2.821	28762656	216.1E6	18.287	21.619
2)	SA Decachlor...	8.814	7.976	75101181	454.9E6	36.986	41.330
Target Compounds							
2)	A alpha-BHC	3.822	3.325	55052065	351.9E6	15.944	20.666 #
3)	MA gamma-BHC...	4.149	3.657	52919198	325.3E6	16.227	20.830 #
4)	MA Heptachlor	4.741	4.005	47896120	328.1E6	17.005	20.901
9)	A Endosulfan I	5.877	5.162	45031397	279.3E6	17.723	21.147
13)	MA Dieldrin	6.148	5.427	100.8E6	604.3E6	34.891	41.677
14)	MA Endrin	6.372	5.703	78071699	549.8E6	35.100	40.837
16)	A 4,4'-DDD	6.511	5.843	62304977	483.4E6	33.166	41.714 #
17)	MA 4,4'-DDT	6.825	6.096	61211093	504.1E6	36.132	41.956
20)	A Methoxychlor	7.300	6.665	166.7E6	1193.4E6	193.462	220.672

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102224\
Data File : PD086182.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2024 18:49
Operator : AR\AJ
Sample : INDA304
Misc :
ALS Vial : 25 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 03:37:29 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102224CLP.M
Quant Title : GC Extractables
QLast Update : Tue Oct 22 16:12:51 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

