

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102225\
 Data File : PD090755.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Oct 2025 14:59
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 23 14:21:53 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 17:43:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.545	2.874	144.8E6	1409.9E6	46.513	47.348
2)	SA Decachlor...	9.068	8.062	216.7E6	1663.4E6	46.324	54.912
Target Compounds							
2)	A alpha-BHC	3.994	3.386	345.5E6	2446.0E6	48.136	47.871
3)	MA gamma-BHC...	4.326	3.722	326.2E6	2338.9E6	47.986	49.501
4)	MA Heptachlor	4.924	4.074	319.0E6	2206.6E6	57.071	49.471
5)	MB Aldrin	5.265	4.360	317.1E6	2230.7E6	48.437	48.160
6)	B beta-BHC	4.513	4.018	123.3E6	942.5E6	47.072	47.528
7)	B delta-BHC	4.761	4.254	327.1E6	2314.0E6	48.379	48.495
8)	B Heptachlo...	5.685	4.864	283.9E6	2037.7E6	49.034	47.791
9)	A Endosulfan I	6.069	5.238	267.3E6	1904.6E6	48.287	48.352
10)	B gamma-Chl...	5.941	5.117	284.7E6	2197.2E6	48.006	48.882
11)	B alpha-Chl...	6.022	5.181	298.1E6	2099.1E6	48.241	48.659
12)	B 4,4'-DDE	6.192	5.366	253.9E6	2150.6E6	47.659	49.058
13)	MA Dieldrin	6.342	5.504	306.7E6	2158.4E6	52.602	48.850
14)	MA Endrin	6.570	5.781	237.7E6	1963.5E6	47.994	48.488
15)	B Endosulfa...	6.783	6.072	244.8E6	1892.2E6	48.956	50.384
16)	A 4,4'-DDD	6.702	5.920	204.9E6	1846.3E6	50.537	50.669
17)	MA 4,4'-DDT	7.018	6.175	209.1E6	1836.9E6	50.858	52.374
18)	B Endrin al...	6.912	6.251	180.4E6	1422.9E6	49.244	51.383
19)	B Endosulfa...	7.146	6.474	223.9E6	1822.2E6	48.268	50.837
20)	A Methoxychlor	7.490	6.745	106.0E6	923.3E6	49.892	52.976
21)	B Endrin ke...	7.627	6.983	240.8E6	2013.4E6	49.559	53.380
22)	Mirex	8.110	7.175	156.4E6	1363.3E6	48.872	53.066

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102225\
Data File : PD090755.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2025 14:59
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 14:21:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

