

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102225\
 Data File : PD090766.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Oct 2025 17:44
 Operator : AR\AJ
 Sample : PB170218BS
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PB170218BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 23 14:23:13 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 17:43:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.874	52634508	513.7E6	16.902	17.251
2)	SA Decachlor...	9.068	8.061	98897682	710.3E6	21.141	23.450
Target Compounds							
2)	A alpha-BHC	3.994	3.386	320.2E6	2278.9E6	44.607	44.600
3)	MA gamma-BHC...	4.325	3.722	300.1E6	2116.4E6	44.135	44.794
4)	MA Heptachlor	4.923	4.074	293.0E6	2048.9E6	52.417	45.936
5)	MB Aldrin	5.264	4.360	291.7E6	2086.2E6	44.560	45.039
6)	B beta-BHC	4.512	4.018	114.1E6	888.1E6	43.546	44.785
7)	B delta-BHC	4.760	4.254	303.6E6	2159.0E6	44.903	45.247
8)	B Heptachlo...	5.685	4.863	259.8E6	1917.3E6	44.876	44.965
9)	A Endosulfan I	6.068	5.238	246.2E6	1750.2E6	44.473	44.432
10)	B gamma-Chl...	5.940	5.116	261.6E6	2056.0E6	44.100	45.740
11)	B alpha-Chl...	6.021	5.181	265.1E6	1963.3E6	42.888	45.513
12)	B 4,4'-DDE	6.191	5.365	233.3E6	2011.4E6	43.798	45.883
13)	MA Dieldrin	6.341	5.503	262.5E6	2036.4E6	45.015	46.090
14)	MA Endrin	6.569	5.780	221.1E6	1865.4E6	44.634	46.066
15)	B Endosulfa...	6.782	6.072	227.9E6	1740.7E6	45.571	46.350
16)	A 4,4'-DDD	6.701	5.920	191.9E6	1703.8E6	47.337	46.758
17)	MA 4,4'-DDT	7.016	6.174	190.5E6	1666.2E6	46.349	47.506
18)	B Endrin al...	6.910	6.250	169.0E6	1312.2E6	46.134	47.383
19)	B Endosulfa...	7.146	6.474	206.0E6	1692.4E6	44.405	47.214
20)	A Methoxychlor	7.490	6.745	97041864	831.1E6	45.654	47.687
21)	B Endrin ke...	7.627	6.983	222.9E6	1833.6E6	45.881	48.613
22)	Mirex	8.109	7.176	152.3E6	1293.2E6	47.581	50.336

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102225\
Data File : PD090766.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2025 17:44
Operator : AR\AJ
Sample : PB170218BS
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PB170218BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 14:23:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

