

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102225\
 Data File : PD090773.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Oct 2025 19:19
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 23 14:24:39 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 17:43:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.544	2.874	148.7E6	1439.9E6	47.761	48.356
2)	SA Decachlor...	9.066	8.061	225.0E6	1689.1E6	48.088	55.761
Target Compounds							
2)	A alpha-BHC	3.994	3.385	354.8E6	2498.0E6	49.429	48.889
3)	MA gamma-BHC...	4.324	3.722	334.3E6	2373.4E6	49.172	50.233
4)	MA Heptachlor	4.923	4.073	323.6E6	2208.8E6	57.892	49.521
5)	MB Aldrin	5.264	4.360	324.5E6	2258.0E6	49.567	48.749
6)	B beta-BHC	4.512	4.018	126.0E6	961.8E6	48.100	48.505
7)	B delta-BHC	4.760	4.254	336.4E6	2360.4E6	49.752	49.467
8)	B Heptachlo...	5.685	4.863	287.4E6	2061.3E6	49.643	48.343
9)	A Endosulfan I	6.068	5.238	272.1E6	1923.6E6	49.153	48.834
10)	B gamma-Chl...	5.939	5.116	292.3E6	2224.2E6	49.275	49.483
11)	B alpha-Chl...	6.021	5.181	297.9E6	2129.6E6	48.203	49.367
12)	B 4,4'-DDE	6.190	5.365	259.6E6	2175.5E6	48.734	49.625
13)	MA Dieldrin	6.341	5.503	310.4E6	2236.2E6	53.222	50.612
14)	MA Endrin	6.568	5.780	242.8E6	2010.9E6	49.010	49.660
15)	B Endosulfa...	6.781	6.072	248.3E6	1922.0E6	49.662	51.179
16)	A 4,4'-DDD	6.700	5.920	210.1E6	1889.4E6	51.818	51.853
17)	MA 4,4'-DDT	7.016	6.174	208.7E6	1822.0E6	50.781	51.949
18)	B Endrin al...	6.910	6.250	181.7E6	1427.6E6	49.604	51.550
19)	B Endosulfa...	7.145	6.473	227.9E6	1861.6E6	49.133	51.934
20)	A Methoxychlor	7.489	6.744	106.2E6	924.3E6	49.956	53.035
21)	B Endrin ke...	7.625	6.982	245.1E6	2028.5E6	50.441	53.781
22)	Mirex	8.108	7.175	159.6E6	1367.1E6	49.856	53.215

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102225\
Data File : PD090773.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Oct 2025 19:19
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 23 14:24:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

