

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102318\
 Data File : PD050004.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 23 Oct 2018 17:20
 Operator : AJ\SJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 24 02:18:27 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101818.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 19 07:43:29 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.232	4.061	11499873	566.6E6	19.182	22.078
28)	SA Decachlor...	9.016	9.175	10848334	279.5E6	17.611	17.687
Target Compounds							
2)	A alpha-BHC	4.632f	0.000	2225307	0	2.633	N.D. #
3)	MA gamma-BHC...	0.000	4.797f	0	63660041	N.D.	1.883 #
6)	B beta-BHC	5.154f	4.950f	2686575	12502090	7.197	0.868 #
8)	B Heptachlo...	0.000	5.940	0	11072551	N.D.	0.418 #
9)	A Endosulfan I	6.287f	0.000	1629143	0	2.305	N.D. #
10)	B gamma-Chl...	0.000	6.184	0	2193134	N.D.	0.084 #
11)	B alpha-Chl...	6.287	0.000	1629143	0	2.076	N.D. #
23)	Chlordane-1	5.154	0.000	2686575	0	101.029	N.D. #
25)	Chlordane-3	0.000	6.184	0	2193134	N.D.	0.605 #
26)	Chlordane-4	6.287	0.000	1629143	0	22.718	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102318\
Data File : PD050004.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 23 Oct 2018 17:20
Operator : AJ\SJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 24 02:18:27 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101818.M
Quant Title : GC Extractables
QLast Update : Fri Oct 19 07:43:29 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 4.232 min
Delta R.T.: 0.007 min
Response: 11499873
Conc: 19.18 ng/ml

Instrument :
ECD_D
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.061 min
Delta R.T.: 0.000 min
Response: 566587002
Conc: 22.08 ng/ml




#2 alpha-BHC

R.T.: 4.632 min
Delta R.T.: -0.046 min
Response: 2225307
Conc: 2.63 ng/ml



#2 alpha-BHC


R.T.: 0.000 min
Exp R.T. : 4.453 min
Response: 0
Conc: N.D.



#3 gamma-BHC (Lindane)


R.T.: 0.000 min
Exp R.T.: 4.970 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#3 gamma-BHC (Lindane)

R.T.: 4.797 min
Delta R.T.: 0.058 min
Response: 63660041
Conc: 1.88 ng/ml




#6 beta-BHC

R.T.: 5.154 min
Delta R.T.: -0.055 min
Response: 2686575
Conc: 7.20 ng/ml



#6 beta-BHC


R.T.: 4.950 min
Delta R.T.: 0.042 min
Response: 12502090
Conc: 0.87 ng/ml



#8 Heptachlor epoxide


R.T.: 0.000 min
Exp R.T.: 5.983 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#8 Heptachlor epoxide

R.T.: 5.940 min
Delta R.T.: -0.008 min
Response: 11072551
Conc: 0.42 ng/ml




#9 Endosulfan I

R.T.: 6.287 min
Delta R.T.: -0.041 min
Response: 1629143
Conc: 2.31 ng/ml



#9 Endosulfan I


R.T.: 0.000 min
Exp R.T.: 6.306 min
Response: 0
Conc: N.D.



#10 gamma-Chlordane


R.T.: 0.000 min
Exp R.T.: 6.207 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#10 gamma-Chlordane

R.T.: 6.184 min
Delta R.T.: 0.001 min
Response: 2193134
Conc: 0.08 ng/ml




#11 alpha-Chlordane

R.T.: 6.287 min
Delta R.T.: 0.021 min
Response: 1629143
Conc: 2.08 ng/ml



#11 alpha-Chlordane


R.T.: 0.000 min
Exp R.T.: 6.255 min
Response: 0
Conc: N.D.



#23 Chlordane-1


R.T.: 5.154 min
Delta R.T.: 0.026 min
Response: 2686575
Conc: 101.03 ng/ml

Instrument :
ECD_D
ClientSampleId :




#23 Chlordane-1

R.T.: 0.000 min
Exp R.T. : 5.063 min
Response: 0
Conc: N.D.




#25 Chlordane-3

R.T.: 0.000 min
Exp R.T. : 6.209 min
Response: 0
Conc: N.D.



#25 Chlordane-3


R.T.: 6.184 min
Delta R.T.: 0.002 min
Response: 2193134
Conc: 0.61 ng/ml



#26 Chlordane-4


R.T.: 6.287 min
Delta R.T.: 0.022 min
Response: 1629143
Conc: 22.72 ng/ml

Instrument :
ECD_D
ClientSampleId :




#26 Chlordane-4

R.T.: 0.000 min
Exp R.T. : 6.257 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 9.016 min
Delta R.T.: 0.010 min
Response: 10848334
Conc: 17.61 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.175 min
Delta R.T.: 0.007 min
Response: 279483128
Conc: 17.69 ng/ml