

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102319\
 Data File : PD055536.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Oct 2019 11:51
 Operator : SG\AJ
 Sample : INDA330
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA330

Manual Integrations
 APPROVED

Ankita
 10/24/2019 12:57:17 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 24 04:16:54 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101819CLP.M
 Quant Title : GC Extractables
 QLast Update : Sat Oct 19 00:42:03 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.290	3.963	18229212	25927656	20.446	19.258
27)	SA Decachlor...	7.966	8.998	32720834	44149746	35.734	33.837
Target Compounds							
2)	A alpha-BHC	3.717	4.349	26221031	37893738	20.879	19.478
3)	MA gamma-BHC...	3.992	4.632	24672684	35442708	20.821	19.221
4)	MA Heptachlor	4.278	5.138	21794676	36017740	23.951	19.314
9)	A Endosulfan I	5.286	6.181	18904725	30314040	19.668	18.495
13)	MA Dieldrin	5.524	6.437	43291549	67473226	39.063m	37.281
14)	MA Endrin	5.778	6.653	37473412	57156236	41.836	40.920m
16)	A 4,4'-DDD	5.913	6.776	34657700	54582476	42.762	38.153
17)	MA 4,4'-DDT	6.148	7.074	29139214	50828476	42.979	37.500
20)	A Methoxychlor	6.697	7.530	83018605	140.8E6	227.654	182.631

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102319\
Data File : PD055536.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Oct 2019 11:51
Operator : SG\AJ
Sample : INDA330
Misc :
ALS Vial : 8 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 24 04:16:54 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101819CLP.M
Quant Title : GC Extractables
QLast Update : Sat Oct 19 00:42:03 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Manual Integrations
APPROVED

Ankita
10/24/2019 12:57:17 PM

Instrument :
ECD_D
ClientSample :
INDA330