

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102418\
 Data File : PD050094.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 Oct 2018 21:12
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Sohil
 10/25/2018 10:27:58 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 25 02:19:55 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101818.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 19 07:43:29 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.232	4.060	30259159	1314.3E6	50.473	51.215
28)	SA Decachlor...	9.017	9.171	28530943	707.2E6	46.317	44.756
Target Compounds							
2)	A alpha-BHC	4.685	4.453	44781853	1982.6E6	52.992	53.168
3)	MA gamma-BHC...	4.978	4.740	42939170	1802.8E6	52.777	53.330
4)	MA Heptachlor	5.291	5.253	39341424	1697.1E6	50.092	52.684
5)	MB Aldrin	5.549	5.561	38221869	1597.2E6	48.931	53.722
6)	B beta-BHC	5.217	4.908	18338610	764.1E6	49.124	53.079
7)	B delta-BHC	5.436	5.128	40214728	1752.6E6	51.360	55.146
8)	B Heptachlo...	5.990	5.948	36385528	1410.2E6	47.889	53.288m
9)	A Endosulfan I	6.336	6.308	33782077	1227.7E6	47.804	52.603
10)	B gamma-Chl...	6.214	6.182	37645123	1380.4E6	48.117	52.584m
11)	B alpha-Chl...	6.273	6.256	37041100	1297.6E6	47.197	51.731m
12)	B 4,4'-DDE	6.417	6.411	29203043	1228.2E6	41.621	52.044 #
13)	MA Dieldrin	6.576	6.565	36140973	1249.4E6	48.156m	51.792
14)	MA Endrin	6.837	6.784	26958359	784.1E6	41.515	40.786
15)	B Endosulfa...	7.100	6.992	32594856	992.9E6	45.772	50.084
16)	A 4,4'-DDD	6.928	6.902	28983882	962.8E6	52.133	53.910
17)	MA 4,4'-DDT	7.168	7.203	24410523	685.9E6	47.070	56.558
18)	B Endrin al...	7.264	7.116	25939088	816.3E6	48.692	51.620
19)	B Endosulfa...	7.470	7.340	29687675	874.5E6	45.944	48.754
20)	A Methoxychlor	7.694	7.659	14011327	346.0E6	47.609	51.688
21)	B Endrin ke...	7.966	7.813	36039468	842.3E6	52.417	50.880

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102418\
Data File : PD050094.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 24 Oct 2018 21:12
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Sohil
10/25/2018 10:27:58 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 25 02:19:55 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101818.M
Quant Title : GC Extractables
QLast Update : Fri Oct 19 07:43:29 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

