

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102721\
 Data File : PD066558.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Oct 2021 13:22
 Operator : AJ\MA
 Sample : INDA301
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA3033

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 27 15:41:34 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102721CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Oct 27 15:40:17 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.440	3.970	38240686	172.5E6	20.000	20.000
27)	SA Decachlor...	8.167	9.024	46238622	209.7E6	40.000	40.000
Target Compounds							
2)	A alpha-BHC	3.877	4.363	57555670	216.4E6	20.000	20.000
3)	MA gamma-BHC...	4.159	4.648	54804249	201.1E6	20.000	20.000
4)	MA Heptachlor	4.453	5.148	56606465	195.2E6	20.000	20.000
9)	A Endosulfan I	5.475	6.197	45558793	156.5E6	20.000	20.000
13)	MA Dieldrin	5.717	6.455	99594530	318.8E6	40.000	40.000
14)	MA Endrin	5.974	6.674	84898975	272.5E6	40.000	40.000
16)	A 4,4'-DDD	6.099	6.799	72665869	259.2E6	40.000	40.000
17)	MA 4,4'-DDT	6.338	7.097	76001421	264.7E6	40.000	40.000
20)	A Methoxychlor	6.885	7.559	187.6E6	645.8E6	200.000	200.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102721\
Data File : PD066558.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2021 13:22
Operator : AJ\MA
Sample : INDA301
Misc :
ALS Vial : 13 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 15:41:34 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102721CLP.M
Quant Title : GC Extractables
QLast Update : Wed Oct 27 15:40:17 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

