

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102820\
 Data File : PD059936.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Oct 2020 17:54
 Operator : DD\AJ
 Sample : INDB401
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 29 02:06:25 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102820CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu Oct 29 02:04:44 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.507	4.025	37754180	102.9E6	38.199	38.916
27)	SA Decachlor...	8.307	9.110	94353698	165.8E6	76.717	78.264
Target Compounds							
5)	MB Aldrin	4.795	5.519	54434674	128.8E6	39.768	40.104
6)	B beta-BHC	4.486	4.868	23805103	61434613	38.058	37.897
7)	B delta-BHC	4.694	5.088	54658320	124.5E6	38.933	38.658
8)	B Heptachlo...	5.241	5.906	49694340	112.1E6	38.691	38.666
10)	B trans-Chl...	5.471	6.140	51745843	118.7E6	39.381	39.854
11)	B cis-Chlor...	5.530	6.214	51752891	117.8E6	38.994	39.370
12)	B 4,4'-DDE	5.694	6.368	104.8E6	231.1E6	80.140	81.410
15)	B Endosulfa...	6.363	6.948	90207001	202.8E6	79.004	79.766
18)	B Endrin al...	6.532	7.072	74996438	168.8E6	77.836	77.994
19)	B Endosulfa...	6.747	7.295	81874003	204.1E6	73.677	78.819
21)	B Endrin ke...	7.242	7.768	102.4E6	213.4E6	78.301	79.630


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102820\
Data File : PD059936.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Oct 2020 17:54
Operator : DD\AJ
Sample : INDB401
Misc :
ALS Vial : 16 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 29 02:06:25 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102820CLP.M
Quant Title : GC Extractables
QLast Update : Thu Oct 29 02:04:44 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.507 min
Delta R.T.: 0.000 min
Response: 37754180
Conc: 38.20 ng/ml

Instrument :
ECD_D
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.025 min
Delta R.T.: 0.000 min
Response: 102866619
Conc: 38.92 ng/ml




#5 Aldrin

R.T.: 4.795 min
Delta R.T.: 0.000 min
Response: 54434674
Conc: 39.77 ng/ml



#5 Aldrin


R.T.: 5.519 min
Delta R.T.: 0.000 min
Response: 128795289
Conc: 40.10 ng/ml



#6 beta-BHC


R.T.: 4.486 min
Delta R.T.: 0.000 min
Response: 23805103
Conc: 38.06 ng/ml

Instrument :
ECD_D
ClientSampleId :




#6 beta-BHC

R.T.: 4.868 min
Delta R.T.: 0.000 min
Response: 61434613
Conc: 37.90 ng/ml




#7 delta-BHC

R.T.: 4.694 min
Delta R.T.: 0.000 min
Response: 54658320
Conc: 38.93 ng/ml



#7 delta-BHC


R.T.: 5.088 min
Delta R.T.: 0.000 min
Response: 124499051
Conc: 38.66 ng/ml



#8 Heptachlor epoxide


R.T.: 5.241 min
Delta R.T.: 0.000 min
Response: 49694340
Conc: 38.69 ng/ml

Instrument :
ECD_D
ClientSampleId :




#8 Heptachlor epoxide

R.T.: 5.906 min
Delta R.T.: 0.000 min
Response: 112114582
Conc: 38.67 ng/ml




#10 trans-Chlordane

R.T.: 5.471 min
Delta R.T.: 0.000 min
Response: 51745843
Conc: 39.38 ng/ml



#10 trans-Chlordane


R.T.: 6.140 min
Delta R.T.: 0.000 min
Response: 118708299
Conc: 39.85 ng/ml



#11 cis-Chlordane


R.T.: 5.530 min
Delta R.T.: 0.000 min
Response: 51752891
Conc: 38.99 ng/ml

Instrument :
ECD_D
ClientSampleId :




#11 cis-Chlordane

R.T.: 6.214 min
Delta R.T.: 0.000 min
Response: 117802159
Conc: 39.37 ng/ml




#12 4,4'-DDE

R.T.: 5.694 min
Delta R.T.: 0.000 min
Response: 104771542
Conc: 80.14 ng/ml



#12 4,4'-DDE


R.T.: 6.368 min
Delta R.T.: 0.000 min
Response: 231122034
Conc: 81.41 ng/ml



#15 Endosulfan II


R.T.: 6.363 min
Delta R.T.: 0.000 min
Response: 90207001
Conc: 79.00 ng/ml

Instrument :
ECD_D
ClientSampleId :




#15 Endosulfan II

R.T.: 6.948 min
Delta R.T.: 0.000 min
Response: 202783894
Conc: 79.77 ng/ml




#18 Endrin aldehyde

R.T.: 6.532 min
Delta R.T.: 0.000 min
Response: 74996438
Conc: 77.84 ng/ml



#18 Endrin aldehyde


R.T.: 7.072 min
Delta R.T.: 0.000 min
Response: 168764006
Conc: 77.99 ng/ml



#19 Endosulfan Sulfate


R.T.: 6.747 min
Delta R.T.: 0.000 min
Response: 81874003
Conc: 73.68 ng/ml

Instrument :
ECD_D
ClientSampleId :




#19 Endosulfan Sulfate

R.T.: 7.295 min
Delta R.T.: 0.000 min
Response: 204075424
Conc: 78.82 ng/ml




#21 Endrin ketone

R.T.: 7.242 min
Delta R.T.: 0.001 min
Response: 102373089
Conc: 78.30 ng/ml



#21 Endrin ketone


R.T.: 7.768 min
Delta R.T.: 0.000 min
Response: 213427785
Conc: 79.63 ng/ml



#27 Decachlorobiphenyl

R.T.: 8.307 min
Delta R.T.: 0.000 min
Response: 94353698
Conc: 76.72 ng/ml

Instrument :
ECD_D
ClientSampleId :



#27 Decachlorobiphenyl

R.T.: 9.110 min
Delta R.T.: 0.000 min
Response: 165835669
Conc: 78.26 ng/ml