

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : DTX102821.M
 Title : GC Extractables
 Last Update : Thu Oct 28 03:24:11 2021
 Response Via : Initial Calibration

Calibration Files

500 =PD066588.D 1000 =PD066586.D 750 =PD066587.D
 250 =PD066589.D 100 =PD066590.D

	Compound	500	1000	750	250	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	1.805	1.786	1.769	1.836	1.854	1.810	E6	1.93
2)	Toxaphene-1	1.562	1.586	1.553	1.746	1.760	1.641	E4	6.25
3)	Toxaphene-2	1.768	1.614	1.616	2.004	2.225	1.845	E4	14.37
4)	Toxaphene-3	3.716	3.636	3.534	4.003	4.205	3.819	E4	7.27
5)	Toxaphene-4	3.833	3.659	3.588	4.185	4.261	3.905	E4	7.80
6)	Toxaphene-5	3.938	3.829	3.790	4.194	3.837	3.917	E4	4.18
7) SA	Decachlorobiphenyl	1.094	1.034	1.036	1.185	1.279	1.126	E6	9.36

Signal #2 Calibration Files

500 =PD066588.D 1000 =PD066586.D 750 =PD066587.D
 250 =PD066589.D 100 =PD066590.D

	Compound	500	1000	750	250	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	7.628	7.113	7.234	8.284	8.873	7.826	E6	9.48
2)	Toxaphene-1	4.407	4.081	4.157	5.042	4.666	4.471	E4	8.80
3)	Toxaphene-2	6.925	6.390	6.235	7.317	7.938	6.961	E4	9.99
4)	Toxaphene-3	1.272	1.196	1.202	1.379	1.387	1.287	E5	7.17
5)	Toxaphene-4	7.414	7.020	6.871	7.785	7.786	7.375	E4	5.75
6)	Toxaphene-5	1.014	0.966	0.959	1.135	1.108	1.036	E5	7.83
7) SA	Decachlorobiphenyl	5.163	4.767	4.785	5.607	5.864	5.237	E6	9.36

(#) = Out of Range