

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102821\
 Data File : PD066592.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Oct 2021 00:12
 Operator : AJ\MA
 Sample : PCHLORICV500
 Misc :
 ALS Vial : 43 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 28 02:01:25 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102821.M
 Quant Title : GC Extractables
 QLast Update : Thu Oct 28 02:00:45 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(Min)
1 SA	Tetrachloro-m-xylene	50.000	52.710	-5.4	103	0.00
23	Chlordane-1	500.000	520.171	-4.0	104	0.00
24	Chlordane-2	500.000	528.241	-5.6	111	0.00
25	Chlordane-3	500.000	530.121	-6.0	104	0.00
26	Chlordane-4	500.000	528.091	-5.6	103	0.00
27	Chlordane-5	500.000	515.456	-3.1	98	0.00
28 SA	Decachlorobiphenyl	50.000	50.793	-1.6	102	0.00

Signal #2

1 SA	Tetrachloro-m-xylene	50.000	51.333	-2.7	104	0.00
23	Chlordane-1	500.000	533.507	-6.7	105	0.00
24	Chlordane-2	500.000	513.294	-2.7	103	0.00
25	Chlordane-3	500.000	508.782	-1.8	104	0.00
26	Chlordane-4	500.000	509.510	-1.9	106	0.00
27	Chlordane-5	500.000	519.578	-3.9	100	0.00
28 SA	Decachlorobiphenyl	50.000	50.707	-1.4	102	0.00

Evaluate Continuing Calibration Report - Not Found

2 MA	alpha-BHC	50.000	0.000	100.0#	0	-3.88#
3 MA	gamma-BHC (Lindane)	50.000	0.000	100.0#	0	-4.16#
4 MA	Heptachlor	50.000	0.000	100.0#	0	-4.45#
5 MB	Aldrin	50.000	0.000	100.0#	0	-4.70#
6 B	beta-BHC	50.000	0.000	100.0#	0	-4.41#
7 B	delta-BHC	50.000	0.000	100.0#	0	-4.61#
8 B	Heptachlor epoxide	50.000	0.000	100.0#	0	-5.14#
9 A	Endosulfan I	50.000	0.000	100.0#	0	-5.47#
10 B	gamma-Chlordane	50.000	0.000	100.0#	0	-5.36#
11 B	alpha-Chlordane	50.000	0.000	100.0#	0	-5.42#
12 B	4,4'-DDE	50.000	0.000	100.0#	0	-5.58#
13 MA	Dieldrin	50.000	0.000	100.0#	0	-5.72#
14 MA	Endrin	50.000	0.000	100.0#	0	-5.97#
15 B	Endosulfan II	50.000	0.000	100.0#	0	-6.25#
16 A	4,4'-DDD	50.000	0.000	100.0#	0	-6.10#
17 MA	4,4'-DDT	50.000	0.000	100.0#	0	-6.34#
18 B	Endrin aldehyde	50.000	0.000	100.0#	0	-6.41#
19 B	Endosulfan Sulfate	50.000	0.000	100.0#	0	-6.63#
20 A	Methoxychlor	50.000	0.000	100.0#	0	-6.88#
21 B	Endrin ketone	50.000	0.000	100.0#	0	-7.12#
22	Mirex	50.000	0.000	100.0#	0	-7.30#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102821\
 Data File : PD066592.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Oct 2021 00:12
 Operator : AJ\MA
 Sample : PCHLORICV500
 Misc :
 ALS Vial : 43 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 28 02:01:25 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102821.M
 Quant Title : GC Extractables
 QLast Update : Thu Oct 28 02:00:45 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(Min)
Signal #2					
2 A alpha-BHC	50.000	0.000	100.0#	0	-4.36#
3 MA gamma-BHC (Lindane)	50.000	0.000	100.0#	0	-4.65#
4 MA Heptachlor	50.000	0.000	100.0#	0	-5.15#
5 MB Aldrin	50.000	0.000	100.0#	0	-5.45#
6 B beta-BHC	50.000	0.000	100.0#	0	-4.82#
7 B delta-BHC	50.000	0.000	100.0#	0	-5.04#
8 B Heptachlor epoxide	50.000	0.000	100.0#	0	-5.84#
9 A Endosulfan I	50.000	0.000	100.0#	0	-6.20#
10 B gamma-Chlordane	50.000	0.000	100.0#	0	-6.07#
11 B alpha-Chlordane	50.000	0.000	100.0#	0	-6.14#
12 B 4,4'-DDE	50.000	0.000	100.0#	0	-6.30#
13 MA Dieldrin	50.000	0.000	100.0#	0	-6.45#
14 MA Endrin	50.000	0.000	100.0#	0	-6.67#
15 B Endosulfan II	50.000	0.000	100.0#	0	-6.89#
16 A 4,4'-DDD	50.000	0.000	100.0#	0	-6.80#
17 MA 4,4'-DDT	50.000	0.000	100.0#	0	-7.10#
18 B Endrin aldehyde	50.000	0.000	100.0#	0	-7.01#
19 B Endosulfan Sulfate	50.000	0.000	100.0#	0	-7.23#
20 A Methoxychlor	50.000	0.000	100.0#	0	-7.56#
21 B Endrin ketone	50.000	0.000	100.0#	0	-7.71#
22 Mirex	50.000	0.000	100.0#	0	-8.16#

(#) = Out of Range

SPCC's out = 0 CCC's out = 0