

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102821\
 Data File : PD066593.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Oct 2021 00:26
 Operator : AJ\MA
 Sample : PTOXICV500
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 28 03:24:59 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX102821.M
 Quant Title : GC Extractables
 QLast Update : Thu Oct 28 03:24:11 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 SA	Tetrachloro-m-xylene	1.810	1.884 E6	-4.1	104	0.00
2	Toxaphene-1	16.413	16.653 E3	-1.5	107	0.00
3	Toxaphene-2	18.455	17.871 E3	3.2	101	0.00
4	Toxaphene-3	38.188	38.550 E3	-0.9	104	0.00
5	Toxaphene-4	39.051	40.010 E3	-2.5	104	0.00
6	Toxaphene-5	39.173	41.072 E3	-4.8	104	0.00
7 SA	Decachlorobiphenyl	1.126	1.139 E6	-1.2	104	0.00

Signal #2

1 SA	Tetrachloro-m-xylene	7.826	7.970 E6	-1.8	104	0.00
2	Toxaphene-1	44.706	45.229 E3	-1.2	103	0.00
3	Toxaphene-2	69.607	69.902 E3	-0.4	101	0.00
4	Toxaphene-3	128.722	134.609 E3	-4.6	106	0.00
5	Toxaphene-4	73.753	76.876 E3	-4.2	104	0.00
6	Toxaphene-5	103.632	108.702 E3	-4.9	107	0.00
7 SA	Decachlorobiphenyl	5.237	5.305 E6	-1.3	103	0.00

Evaluate Continuing Calibration Report - Not Founds

Signal #2

(#) = Out of Range

SPCC's out = 0 CCC's out = 0