

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102821\
 Data File : PD066576.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Oct 2021 20:31
 Operator : AJ\MA
 Sample : PSTDICC100
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC100

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 28 01:28:18 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102821.M
 Quant Title : GC Extractables
 QLast Update : Thu Oct 28 01:27:03 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.440	3.970	182.2E6	728.3E6	97.428	92.101
2)	SA Decachlor...	8.166	9.023	103.6E6	478.3E6	92.867	90.951
Target Compounds							
2)	A alpha-BHC	3.876	4.363	315.5E6	1021.9E6	101.257	94.864
3)	MA gamma-BHC...	4.158	4.647	288.4E6	924.1E6	98.842	94.792
4)	MA Heptachlor	4.452	5.148	300.2E6	883.4E6	100.082	93.576
5)	MB Aldrin	4.697	5.451	293.0E6	830.8E6	100.233	94.178
6)	B beta-BHC	4.407	4.825	115.7E6	402.8E6	95.782	90.993
7)	B delta-BHC	4.608	5.039	276.4E6	867.5E6	100.372	95.054
8)	B Heptachlo...	5.138	5.839	257.1E6	693.9E6	99.198	97.681
9)	A Endosulfan I	5.474	6.196	236.1E6	711.9E6	99.250	93.032
10)	B gamma-Chl...	5.362	6.073	261.9E6	758.8E6	100.168	93.739
11)	B alpha-Chl...	5.420	6.144	254.7E6	747.7E6	99.475	93.231
12)	B 4,4'-DDE	5.583	6.305	237.3E6	730.3E6	100.920	94.541
13)	MA Dieldrin	5.716	6.454	253.3E6	761.3E6	100.467	94.422
14)	MA Endrin	5.973	6.673	208.7E6	626.6E6	99.676	93.965
15)	B Endosulfa...	6.246	6.887	192.7E6	635.5E6	98.240	92.558
16)	A 4,4'-DDD	6.098	6.799	188.9E6	624.3E6	99.847	93.985
17)	MA 4,4'-DDT	6.337	7.096	196.6E6	672.3E6	100.456	94.593
18)	B Endrin al...	6.414	7.012	146.5E6	508.3E6	96.694	92.328
19)	B Endosulfa...	6.625	7.234	179.1E6	648.2E6	97.558	92.321
20)	A Methoxychlor	6.885	7.558	95242541	357.2E6	94.329	90.544
21)	B Endrin ke...	7.118	7.707	206.9E6	750.7E6	97.152	92.678
22)	Mirex	7.298	8.155	141.3E6	509.0E6	94.522	90.136

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102821\
Data File : PD066576.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2021 20:31
Operator : AJ\MA
Sample : PSTDICC100
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC100

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 28 01:28:18 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102821.M
Quant Title : GC Extractables
QLast Update : Thu Oct 28 01:27:03 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

