

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102825\
 Data File : PD090876.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Oct 2025 12:21
 Operator : AR\AJ
 Sample : Q3447-04
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 P001-TW2-251023-01

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 10/29/2025
 Supervised By : Sohil Jodhani 10/29/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 29 01:42:42 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 17 17:43:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.544	2.873	54341002	531.2E6	17.450	17.838
28) SA Decachlor...	9.060	8.058	75680242	512.5E6	16.178m	16.918

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102825\
Data File : PD090876.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Oct 2025 12:21
Operator : AR\AJ
Sample : Q3447-04
Misc :
ALS Vial : 27 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
P001-TW2-251023-01


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 10/29/2025
Supervised By : Sohil Jodhani 10/29/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 29 01:42:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD101725.M
Quant Title : GC Extractables
QLast Update : Fri Oct 17 17:43:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.544 min
Delta R.T.: -0.001 min
Response: 54341002
Conc: 17.45 ng/ml

Instrument :
ECD_D
ClientSampleId :
P001-TW2-251023-01


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/29/2025
Supervised By :Sohil Jodhani 10/29/2025




#1 Tetrachloro-m-xylene

R.T.: 2.873 min
Delta R.T.: -0.001 min
Response: 531168947
Conc: 17.84 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.060 min
Delta R.T.: -0.004 min
Response: 75680242
Conc: 16.18 ng/ml m



#28 Decachlorobiphenyl

R.T.: 8.058 min
Delta R.T.: -0.004 min
Response: 512477592
Conc: 16.92 ng/ml