

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102920\
 Data File : PD060007.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Oct 2020 03:01
 Operator : DD\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 30 09:25:29 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102920.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 30 04:34:44 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.507	4.025	41646294	136.6E6	44.782	47.209
28)	SA Decachlor...	8.308	9.113	48500702	111.4E6	41.034	48.067
Target Compounds							
2)	A alpha-BHC	3.949	4.416	67455553	221.0E6	45.584	47.716
3)	MA gamma-BHC...	4.236	4.702	64092048	203.1E6	45.353	47.919
4)	MA Heptachlor	4.544	5.215	63926089	193.9E6	43.693	47.367
5)	MB Aldrin	4.796	5.521	62334158	178.2E6	44.865	46.915
6)	B beta-BHC	4.486	4.870	26641591	81882672	44.060	47.097
7)	B delta-BHC	4.695	5.089	62316265	169.8E6	44.936	55.448
8)	B Heptachlo...	5.243	5.908	58375596	157.4E6	44.495	51.279
9)	A Endosulfan I	5.585	6.266	55144816	157.0E6	44.385	48.272
10)	B gamma-Chl...	5.472	6.142	59597162	172.7E6	44.325	48.882
11)	B alpha-Chl...	5.531	6.215	59333715	169.4E6	44.322	48.078
12)	B 4,4'-DDE	5.695	6.370	59366387	162.1E6	44.646	48.288
13)	MA Dieldrin	5.831	6.523	59866871	169.2E6	44.883	48.275
14)	MA Endrin	6.090	6.742	50795128	137.1E6	42.712	45.261
15)	B Endosulfa...	6.363	6.950	52392014	144.8E6	43.549	47.389
16)	A 4,4'-DDD	6.215	6.861	49537743	136.0E6	45.140	48.064
17)	MA 4,4'-DDT	6.458	7.161	51037967	138.3E6	44.695	47.276
18)	B Endrin al...	6.533	7.074	43483221	121.8E6	45.264	48.205
19)	B Endosulfa...	6.747	7.298	45336349	136.6E6	42.485	47.783
20)	A Methoxychlor	7.007	7.618	27448987	74030816	45.289	47.125
21)	B Endrin ke...	7.243	7.770	59278091	159.4E6	46.237	51.245
22)	Mirex	7.438	8.226	41242406	96158634	40.311	46.703

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102920\
Data File : PD060007.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Oct 2020 03:01
Operator : DD\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 30 09:25:29 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102920.M
Quant Title : GC Extractables
QLast Update : Fri Oct 30 04:34:44 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

