

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102920\
 Data File : PD060002.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Oct 2020 01:50
 Operator : DD\AJ
 Sample : PB132706BL
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PB132706BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 30 09:01:41 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX102920.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 30 08:43:56 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.508	4.025	17887622	57653539	19.619	19.992
7) SA Decachlor...	8.308	9.114	22602233	50534715	19.680	21.518

Target Compounds


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102920\
Data File : PD060002.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Oct 2020 01:50
Operator : DD\AJ
Sample : PB132706BL
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PB132706BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 30 09:01:41 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX102920.M
Quant Title : GC Extractables
QLast Update : Fri Oct 30 08:43:56 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.508 min
Delta R.T.: 0.000 min
Response: 17887622
Conc: 19.62 ng/ml

Instrument :
ECD_D
ClientSampleId :
PB132706BL




#1 Tetrachloro-m-xylene

R.T.: 4.025 min
Delta R.T.: 0.000 min
Response: 57653539
Conc: 19.99 ng/ml



#7 Decachlorobiphenyl

R.T.: 8.308 min
Delta R.T.: -0.001 min
Response: 22602233
Conc: 19.68 ng/ml



#7 Decachlorobiphenyl

R.T.: 9.114 min
Delta R.T.: 0.000 min
Response: 50534715
Conc: 21.52 ng/ml