

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102921\
 Data File : PD066618.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Oct 2021 18:08
 Operator : AJ\MA
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 29 02:04:57 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102821.M
 Quant Title : GC Extractables
 QLast Update : Thu Oct 28 02:17:57 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.439	3.969	91686672	390.6E6	49.140	48.272
2)	SA Decachlor...	8.164	9.022	55314132	262.0E6	47.874	48.634
Target Compounds							
2)	A alpha-BHC	3.876	4.362	153.5E6	534.4E6	50.198	48.439
3)	MA gamma-BHC...	4.158	4.647	143.0E6	483.5E6	47.350	48.517
4)	MA Heptachlor	4.451	5.146	148.0E6	466.8E6	49.901	47.619
5)	MB Aldrin	4.696	5.450	145.0E6	439.8E6	50.300	48.451
6)	B beta-BHC	4.406	4.823	59366458	220.5E6	47.526	47.350
7)	B delta-BHC	4.608	5.039	135.5E6	453.2E6	50.416	49.586
8)	B Heptachlo...	5.137	5.838	129.1E6	369.8E6	49.896	51.022
9)	A Endosulfan I	5.474	6.195	118.3E6	381.1E6	49.830	48.328
10)	B gamma-Chl...	5.361	6.072	130.4E6	405.2E6	50.060	48.562
11)	B alpha-Chl...	5.420	6.144	127.7E6	400.4E6	49.770	48.269
12)	B 4,4'-DDE	5.583	6.304	117.7E6	385.0E6	50.524	48.852
13)	MA Dieldrin	5.716	6.453	125.4E6	403.3E6	50.111	48.608
14)	MA Endrin	5.972	6.672	99362264	317.6E6	47.477	46.149
15)	B Endosulfa...	6.245	6.886	97753529	343.0E6	49.158	48.580
16)	A 4,4'-DDD	6.098	6.798	94629553	333.3E6	50.473	49.043
17)	MA 4,4'-DDT	6.336	7.095	97354152	354.3E6	50.614	48.854
18)	B Endrin al...	6.414	7.011	76313535	277.9E6	49.370	48.860
19)	B Endosulfa...	6.626	7.233	91655148	351.8E6	49.247	48.494
20)	A Methoxychlor	6.884	7.557	50129668	197.7E6	48.543	48.520
21)	B Endrin ke...	7.117	7.706	107.8E6	410.5E6	49.410	48.817
22)	Mirex	7.299	8.154	74560439	281.8E6	47.931	47.649

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD102921\
Data File : PD066618.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Oct 2021 18:08
Operator : AJ\MA
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 29 02:04:57 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102821.M
Quant Title : GC Extractables
QLast Update : Thu Oct 28 02:17:57 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

