

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD110119.M
 Title : GC Extractables
 Last Update : Thu Oct 31 06:44:18 2019
 Response Via : Initial Calibration

Calibration Files

50 =PD055724.D 100 =PD055722.D 75 =PD055723.D
 25 =PD055725.D 5 =PD055726.D

	Compound	50	100	75	25	5	Avg	%RSD
1) SA	Tetrachloro-m-xylene	0.898	0.922	0.891	0.880	1.026	0.923 E6	6.44
2) MA	alpha-BHC						0.000	-1.00
3) MA	gamma-BHC (Lindane)						0.000	-1.00
4) MA	Heptachlor						0.000	-1.00
5) MB	Aldrin						0.000	-1.00
6) B	beta-BHC						0.000	-1.00
7) B	delta-BHC						0.000	-1.00
8) B	Heptachlor epoxide						0.000	-1.00
9) A	Endosulfan I						0.000	-1.00
10) B	gamma-Chlordane						0.000	-1.00
11) B	alpha-Chlordane						0.000	-1.00
12) B	4,4'-DDE						0.000	-1.00
13) MA	Dieldrin						0.000	-1.00
14) MA	Endrin						0.000	-1.00
15) B	Endosulfan II						0.000	-1.00
16) A	4,4'-DDD						0.000	-1.00
17) MA	4,4'-DDT						0.000	-1.00
18) B	Endrin aldehyde						0.000	-1.00
19) B	Endosulfan Sulfate						0.000	-1.00
20) A	Methoxychlor						0.000	-1.00
21) B	Endrin ketone						0.000	-1.00
22)	Mirex						0.000	-1.00
23)	Chlordane-1	2.645	2.666	2.579	2.493	2.583	2.593 E4	2.62
24)	Chlordane-2	3.059	3.105	3.047	3.126	3.565	3.180 E4	6.83
25)	Chlordane-3	0.949	0.979	0.933	0.911	1.036	0.962 E5	5.03
26)	Chlordane-4	8.311	8.368	8.242	7.892	9.237	8.410 E4	5.92
27)	Chlordane-5	3.078	3.079	3.008	2.960	3.057	3.036 E4	1.69
28) SA	Decachlorobiphenyl	0.909	0.875	0.865	0.931	1.173	0.950 E6	13.36

Signal #2 Calibration Files

50 =PD055724.D 100 =PD055722.D 75 =PD055723.D
 25 =PD055725.D 5 =PD055726.D

	Compound	50	100	75	25	5	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.156	1.161	1.128	1.157	1.379	1.196 E6	8.60
2) A	alpha-BHC						0.000	-1.00
3) MA	gamma-BHC (Lindane)						0.000	-1.00
4) MA	Heptachlor						0.000	-1.00
5) MB	Aldrin						0.000	-1.00
6) B	beta-BHC						0.000	-1.00
7) B	delta-BHC						0.000	-1.00
8) B	Heptachlor epoxide						0.000	-1.00
9) A	Endosulfan I						0.000	-1.00
10) B	gamma-Chlordane						0.000	-1.00
11) B	alpha-Chlordane						0.000	-1.00
12) B	4,4'-DDE						0.000	-1.00
13) MA	Dieldrin						0.000	-1.00
14) MA	Endrin						0.000	-1.00
15) B	Endosulfan II						0.000	-1.00
16) A	4,4'-DDD						0.000	-1.00
17) MA	4,4'-DDT						0.000	-1.00
18) B	Endrin aldehyde						0.000	-1.00
19) B	Endosulfan Sulfate						0.000	-1.00
20) A	Methoxychlor						0.000	-1.00
21) B	Endrin ketone						0.000	-1.00
22)	Mirex						0.000	-1.00
23)	Chlordane-1	4.524	4.550	4.408	4.533	5.188	4.641 E4	6.70

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD110119.M
 Title : GC Extractables
 Last Update : Thu Oct 31 06:44:18 2019
 Response Via : Initial Calibration

Calibration Files

50 =PD055724.D 100 =PD055722.D 75 =PD055723.D
 25 =PD055725.D 5 =PD055726.D

	Compound	50	100	75	25	5	Avg	%RSD
24)	Chlordane-2	4.988	4.693	4.680	5.138	6.766	5.253	E4 16.53
25)	Chlordane-3	2.138	2.182	2.093	2.082	2.455	2.190	E5 7.00
26)	Chlordane-4	2.545	2.597	2.471	2.543	2.996	2.630	E5 7.96
27)	Chlordane-5	4.596	4.656	4.548	4.587	4.923	4.662	E4 3.23
28) SA	Decachlorobiphenyl	1.316	1.238	1.247	1.370	1.780	1.390	E6 16.14

(#) = Out of Range