

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD110119\
 Data File : PD055734.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Oct 2019 01:57
 Operator : SG\AJ
 Sample : PTOXICV500
 Misc :
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 ICVPD110119TOX

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 31 07:16:17 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX110119.M
 Quant Title : GC Extractables
 QLast Update : Thu Oct 31 07:14:24 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 SA Tetrachloro-m-xylene	848.358	798.447 E3	5.9	99	0.00
2 Toxaphene-1	6.569	5.914 E3	10.0	92	0.00
3 Toxaphene-2	16.667	16.180 E3	2.9	95	0.00
4 Toxaphene-3	16.395	16.175 E3	1.3	98	0.00
5 Toxaphene-4	12.557	12.110 E3	3.6	93	0.00
6 Toxaphene-5	20.629	19.879 E3	3.6	96	0.00
7 SA Decachlorobiphenyl	942.531	863.483 E3	8.4	96	0.00

Signal #2

1 SA Tetrachloro-m-xylene	1.196	1.128 E6	5.7	99	0.00
2 Toxaphene-1	8.622	7.975 E3	7.5	100	0.00
3 Toxaphene-2	10.008	9.788 E3	2.2	103	0.00
4 Toxaphene-3	42.305	39.244 E3	7.2	98	0.00
5 Toxaphene-4	25.733	24.110 E3	6.3	97	0.00
6 Toxaphene-5	25.550	23.768 E3	7.0	98	0.00
7 SA Decachlorobiphenyl	1.391	1.267 E6	8.9	96	0.00

Evaluate Continuing Calibration Report - Not Founds

Signal #2

(#) = Out of Range

SPCC's out = 0 CCC's out = 0