

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD110124\
Data File : PD086302.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Nov 2024 23:37
Operator : AR\AJ
Sample : INDA307
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :

ECD_D

LabSampleId :

INDA307


Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 11/04/2024

Supervised By :Ankita Jodhani 11/04/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 00:25:28 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102524CLP.M
Quant Title : GC Extractables
QLast Update : Fri Oct 25 14:36:38 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(9) Endosulfan I (A)

6.091min 22.420 ng/ml

response 45412686

(9) Endosulfan I #2 (A)

5.264min 22.649 ng/ml

response 260050020

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD110124\
Data File : PD086302.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Nov 2024 23:37
Operator : AR\AJ
Sample : INDA307
Misc :
ALS Vial : 26 Sample Multiplier: 1



Instrument :
ECD_D
LabSampleId :
INDA307

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 11/04/2024
Supervised By :Ankita Jodhani 11/04/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 00:25:28 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102524CLP.M
Quant Title : GC Extractables
QLast Update : Fri Oct 25 14:36:38 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(9) Endosulfan I (A)
6.090min 22.600 ng/ml m
response 45777409

(9) Endosulfan I #2 (A)
5.263min 22.631 ng/ml m
response 259834438

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD110124\
Data File : PD086302.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Nov 2024 23:37
Operator : AR\AJ
Sample : INDA307
Misc :
ALS Vial : 26 Sample Multiplier: 1



Instrument :
ECD_D
LabSampleId :
INDA307

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 11/04/2024
Supervised By :Ankita Jodhani 11/04/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 00:25:28 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102524CLP.M
Quant Title : GC Extractables
QLast Update : Fri Oct 25 14:36:38 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(17) 4,4'-DDT (MA)

7.037min 45.404 ng/ml

response 75698638

(17) 4,4'-DDT #2 (MA)

6.204min 44.710 ng/ml

response 458802132

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD110124\
Data File : PD086302.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Nov 2024 23:37
Operator : AR\AJ
Sample : INDA307
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :

ECD_D

LabSampleId :

INDA307


Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 11/04/2024

Supervised By :Ankita Jodhani 11/04/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 04 00:25:28 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102524CLP.M
Quant Title : GC Extractables
QLast Update : Fri Oct 25 14:36:38 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(17) 4,4'-DDT (MA)

7.036min 44.911 ng/ml m

response 74877613

(17) 4,4'-DDT #2 (MA)

6.204min 44.710 ng/ml

response 458802132