

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD110222\
 Data File : PD072772.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Nov 2022 11:45
 Operator : AR\AJ
 Sample : N5321-23DL 4X
 Misc :
 ALS Vial : 10 Sample Multiplier: 1


Instrument :
 ECD_D
 ClientSampleId :
 DBWW5DL

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 11/03/2022
 Supervised By :Ankita Jodhani 11/04/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 02 22:25:00 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102622CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Oct 26 16:50:44 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(12) 4,4'-DDE (B)

5.163min 1.053 ng/ml

response 2865478

(12) 4,4'-DDE #2 (B)

6.089min 0.272 ng/ml

response 4852630

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD110222\
Data File : PD072772.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Nov 2022 11:45
Operator : AR\AJ
Sample : N5321-23DL 4X
Misc :
ALS Vial : 10 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
DBWW5DL

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 11/03/2022
Supervised By :Ankita Jodhani 11/04/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 02 22:25:00 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102622CLP.M
Quant Title : GC Extractables
QLast Update : Wed Oct 26 16:50:44 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(12) 4,4'-DDE (B)

5.163min 1.053 ng/ml
response 2865478

(12) 4,4'-DDE #2 (B)

6.087min 0.315 ng/ml m
response 5620624