

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Data\PD110322\  
Data File : PD072806.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 03 Nov 2022 14:37  
Operator : AR\AJ  
Sample : PIBLK391  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

## Instrument :

ECD\_D

## LabSampleId :


PIBLK391

## Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 11/04/2022  
Supervised By :Ankita Jodhani 11/04/2022

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Nov 04 01:20:32 2022  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Method\PD102622CLP.M  
Quant Title : GC Extractables  
QLast Update : Wed Oct 26 16:50:44 2022  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 µl  
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1  
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(1) Tetrachloro-m-xylene (SA)

2.710min 19.989 ng/ml

response 46473577

(1) Tetrachloro-m-xylene #2 (SA)

3.461min 22.548 ng/ml

response 903553912

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Data\PD110322\  
Data File : PD072806.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 03 Nov 2022 14:37  
Operator : AR\AJ  
Sample : PIBLK391  
Misc :  
ALS Vial : 2 Sample Multiplier: 1


Instrument :  
ECD\_D  
LabSampleId :  
PIBLK391

## Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 11/04/2022  
Supervised By :Ankita Jodhani 11/04/2022

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Nov 04 01:20:32 2022  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Method\PD102622CLP.M  
Quant Title : GC Extractables  
QLast Update : Wed Oct 26 16:50:44 2022  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 µl  
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1  
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(1) Tetrachloro-m-xylene (SA)

2.709min 20.692 ng/ml m

response 48106939

(1) Tetrachloro-m-xylene #2 (SA)

3.461min 22.548 ng/ml

response 903553912