

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD110421\
 Data File : PD066722.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2021 11:42
 Operator : AR\AJ
 Sample : PIBLK138
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 08 04:55:31 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102721CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Oct 27 16:50:52 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.438	3.968	39281949	175.1E6	20.321	20.219
27)	SA Decachlor...	8.162	9.021	44364617	229.4E6	37.864	43.995
Target Compounds							
2)	A alpha-BHC	0.000	4.349	0	399013	N.D.	<MDL
4)	MA Heptachlor	4.413f	5.158	1225115	785866	0.431	<MDL #
6)	B beta-BHC	4.413	4.842	1225115	2439229	0.975	0.495 #
9)	A Endosulfan I	0.000	6.216	0	253052	N.D.	<MDL
10)	B trans-Chl...	0.000	6.083	0	123747	N.D.	<MDL
11)	B cis-Chlor...	0.000	6.156	0	708095	N.D.	<MDL
12)	B 4,4'-DDE	0.000	6.301	0	124683	N.D.	<MDL
13)	MA Dieldrin	0.000	6.437	0	113223	N.D.	<MDL
14)	MA Endrin	0.000	6.630f	0	90437	N.D.	<MDL
15)	B Endosulfa...	6.188f	6.919	19579005	252803	9.429	<MDL #
16)	A 4,4'-DDD	0.000	6.799	0	49646	N.D.	<MDL
17)	MA 4,4'-DDT	0.000	7.116	0	922608	N.D.	0.140
18)	B Endrin al...	0.000	7.001	0	164089	N.D.	<MDL
19)	B Endosulfa...	0.000	7.230	0	629632	N.D.	<MDL
20)	A Methoxychlor	0.000	7.542	0	482102	N.D.	0.149
22)	Toxaphene-1	0.000	6.353	0	34298	N.D.	0.848
23)	Toxaphene-2	6.188	6.533	19579005	-454743	1217.752	N.D. #
24)	Toxaphene-3	0.000	7.116	0	922608	N.D.	8.108
25)	Toxaphene-4	0.000	7.213	0	96018	N.D.	0.849
26)	Toxaphene-5	0.000	7.643	0	28920	N.D.	0.340


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD110421\
Data File : PD066722.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2021 11:42
Operator : AR\AJ
Sample : PIBLK138
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 08 04:55:31 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102721CLP.M
Quant Title : GC Extractables
QLast Update : Wed Oct 27 16:50:52 2021
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene

R.T.: 3.438 min
Delta R.T.: -0.002 min
Response: 39281949
Conc: 20.32 ng/ml




#1 Tetrachloro-m-xylene

R.T.: 3.968 min
Delta R.T.: -0.002 min
Response: 175107966
Conc: 20.22 ng/ml




#4 Heptachlor

R.T.: 4.413 min
Delta R.T.: -0.039 min
Response: 1225115
Conc: 0.43 ng/ml




#4 Heptachlor

R.T.: 5.158 min
Delta R.T.: 0.010 min
Response: 785866
Conc: N.D.




#6 beta-BHC

R.T.: 4.413 min
Delta R.T.: 0.006 min
Response: 1225115
Conc: 0.98 ng/ml




#6 beta-BHC

R.T.: 4.842 min
Delta R.T.: 0.018 min
Response: 2439229
Conc: 0.50 ng/ml




#15 Endosulfan II

R.T.: 6.188 min
Delta R.T.: -0.059 min
Response: 19579005
Conc: 9.43 ng/ml




#15 Endosulfan II

R.T.: 6.919 min
Delta R.T.: 0.032 min
Response: 252803
Conc: N.D.




#17 4,4'-DDT

R.T.: 0.000 min
Exp R.T. : 6.338 min
Response: 0
Conc: N.D.




#17 4,4'-DDT

R.T.: 7.116 min
Delta R.T.: 0.019 min
Response: 922608
Conc: 0.14 ng/ml




#20 Methoxychlor

R.T.: 0.000 min
Exp R.T. : 6.886 min
Response: 0
Conc: N.D.




#20 Methoxychlor

R.T.: 7.542 min
Delta R.T.: -0.016 min
Response: 482102
Conc: 0.15 ng/ml




#22 Toxaphene-1

R.T.: 0.000 min
Exp R.T.: 5.679 min
Response: 0
Conc: N.D.




#22 Toxaphene-1

R.T.: 6.353 min
Delta R.T.: 0.000 min
Response: 34298
Conc: 0.85 ng/ml




#23 Toxaphene-2

R.T.: 6.188 min
Delta R.T.: 0.028 min
Response: 19579005
Conc: 1217.75 ng/ml




#23 Toxaphene-2

R.T.: 6.533 min
Delta R.T.: -0.012 min
Response: -454743
Conc: N.D.




#24 Toxaphene-3

R.T.: 0.000 min
Exp R.T. : 6.280 min
Response: 0
Conc: N.D.




#24 Toxaphene-3

R.T.: 7.116 min
Delta R.T.: -0.013 min
Response: 922608
Conc: 8.11 ng/ml




#25 Toxaphene-4

R.T.: 0.000 min
Exp R.T. : 6.525 min
Response: 0
Conc: N.D.




#25 Toxaphene-4

R.T.: 7.213 min
Delta R.T.: -0.002 min
Response: 96018
Conc: 0.85 ng/ml




#26 Toxaphene-5

R.T.: 0.000 min
Exp R.T. : 7.010 min
Response: 0
Conc: N.D.




#26 Toxaphene-5

R.T.: 7.643 min
Delta R.T.: 0.010 min
Response: 28920
Conc: 0.34 ng/ml



#27 Decachlorobiphenyl

R.T.: 8.162 min
Delta R.T.: -0.006 min
Response: 44364617
Conc: 37.86 ng/ml



#27 Decachlorobiphenyl

R.T.: 9.021 min
Delta R.T.: -0.002 min
Response: 229447632
Conc: 44.00 ng/ml