

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Data\PD110521\  
 Data File : PD066748.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 05 Nov 2021 14:09  
 Operator : AR\AJ  
 Sample : M4412-20  
 Misc :  
 ALS Vial : 12 Sample Multiplier: 1

Instrument :  
 ECD\_D  
 ClientSampleId :  
 BG374

Manual Integrations  
 APPROVED

Reviewed By :Abdul Mirza 11/08/2021  
 Supervised By :mohammad ahmed 11/08/2021

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Nov 08 05:40:45 2021  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Method\PD102721CLP.M  
 Quant Title : GC Extractables  
 QLast Update : Wed Oct 27 16:50:52 2021  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 1 µl  
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1  
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm


|                             | Compound        | RT#1  | RT#2  | Resp#1   | Resp#2  | ng/ml  | ng/ml  |
|-----------------------------|-----------------|-------|-------|----------|---------|--------|--------|
| -----                       |                 |       |       |          |         |        |        |
| System Monitoring Compounds |                 |       |       |          |         |        |        |
| 1)                          | SA Tetrachlo... | 3.439 | 3.968 | 37231357 | 163.0E6 | 19.260 | 18.820 |
| 27)                         | SA Decachlor... | 8.163 | 9.023 | 37952891 | 197.5E6 | 32.392 | 37.862 |
| Target Compounds            |                 |       |       |          |         |        |        |
| 20)                         | A Methoxychlor  | 6.859 | 7.546 | 1282393  | 4552529 | 1.368m | 1.409m |
| -----                       |                 |       |       |          |         |        |        |

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Data\PD110521\  
Data File : PD066748.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 05 Nov 2021 14:09  
Operator : AR\AJ  
Sample : M4412-20  
Misc :  
ALS Vial : 12 Sample Multiplier: 1

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Nov 08 05:40:45 2021  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Method\PD102721CLP.M  
Quant Title : GC Extractables  
QLast Update : Wed Oct 27 16:50:52 2021  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 µl  
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1  
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Instrument :  
ECD\_D  
ClientSampled :  
BG374

Manual Integrations  
APPROVED

Reviewed By : Abdul Miza 11/08/2021  
Supervised By : mohammad ahmed 11/08/2021