

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD110618CLP.M
 Title : GC Extractables
 Last Update : Tue Nov 06 03:46:44 2018
 Response Via : Initial Calibration

Calibration Files

3 =PD050204.D 1 =PD050200.D 2 =PD050202.D
 4 =PD050206.D 5 =PD050208.D

	Compound	3	1	2	4	5	Avg		%RSD
1)	SA Tetrachloro-m-xylene	6.077	6.116	6.200	6.058	5.958	6.082	E5	1.45
2)	A alpha-BHC						0.000		-1.00
3)	MA gamma-BHC (Lindane)						0.000		-1.00
4)	MA Heptachlor						0.000		-1.00
5)	MB Aldrin	7.780	7.596	7.684	7.841	7.991	7.778	E5	1.94
6)	B beta-BHC	3.953	4.018	4.029	3.799	3.687	3.897	E5	3.82
7)	B delta-BHC	8.425	7.917	8.237	8.638	8.878	8.419	E5	4.38
8)	B Heptachlor epoxide	7.562	7.547	7.660	7.527	7.539	7.567	E5	0.70
9)	A Endosulfan I						0.000		-1.00
10)	B trans-Chlordane	7.995	8.729	8.374	7.782	7.778	8.132	E5	5.08
11)	B cis-Chlordane	7.621	8.301	7.927	7.484	7.404	7.747	E5	4.75
12)	B 4,4'-DDE	7.387	6.887	7.159	7.558	7.645	7.327	E5	4.20
13)	MA Dieldrin						0.000		-1.00
14)	MA Endrin						0.000		-1.00
15)	B Endosulfan II	7.152	7.095	7.093	7.098	7.146	7.117	E5	0.41
16)	A 4,4'-DDD						0.000		-1.00
17)	MA 4,4'-DDT						0.000		-1.00
18)	B Endrin aldehyde	5.558	5.529	5.601	5.486	5.458	5.526	E5	1.02
19)	B Endosulfan Sulfate	6.661	6.560	6.647	6.599	6.608	6.615	E5	0.61
20)	A Methoxychlor						0.000		-1.00
21)	B Endrin ketone	7.361	7.100	7.248	7.346	7.356	7.282	E5	1.54
22)	Toxaphene-1						0.000		-1.00
23)	Toxaphene-2						0.000		-1.00
24)	Toxaphene-3						0.000		-1.00
25)	Toxaphene-4						0.000		-1.00
26)	Toxaphene-5						0.000		-1.00
27)	SA Decachlorobiphenyl	6.392	6.774	6.610	6.159	5.914	6.370	E5	5.40

Signal #2 Calibration Files

3 =PD050204.D 1 =PD050200.D 2 =PD050202.D
 4 =PD050206.D 5 =PD050208.D

Compound			3	1	2	4	5	Avg		%RSD
1) SA	Tetrachloro-m-xylene		2.723	2.896	2.864	2.602	2.384	2.694	E7	7.77
2) A	alpha-BHC							0.000		-1.00
3) MA	gamma-BHC (Lindane)							0.000		-1.00
4) MA	Heptachlor							0.000		-1.00
5) MB	Aldrin		3.519	3.758	3.697	3.330	3.099	3.480	E7	7.78
6) B	beta-BHC		1.799	2.040	1.968	1.663	1.505	1.795	E7	12.19
7) B	delta-BHC		3.849	3.989	3.963	3.700	3.487	3.798	E7	5.46
8) B	Heptachlor epoxide		3.231	3.566	3.453	3.016	2.799	3.213	E7	9.74
9) A	Endosulfan I							0.000		-1.00
10) B	trans-Chlordane		3.217	3.406	3.376	3.032	2.844	3.175	E7	7.48
11) B	cis-Chlordane		3.073	3.279	3.233	2.885	2.704	3.035	E7	7.93
12) B	4,4'-DDE		2.875	3.035	3.001	2.745	2.601	2.851	E7	6.33
13) MA	Dieldrin							0.000		-1.00
14) MA	Endrin							0.000		-1.00
15) B	Endosulfan II		2.449	2.647	2.588	2.302	2.150	2.427	E7	8.43
16) A	4,4'-DDD							0.000		-1.00
17) MA	4,4'-DDT							0.000		-1.00
18) B	Endrin aldehyde		1.954	2.111	2.073	1.820	1.696	1.931	E7	8.99
19) B	Endosulfan Sulfate		2.276	2.491	2.419	2.133	1.989	2.262	E7	9.06
20) A	Methoxychlor							0.000		-1.00
21) B	Endrin ketone		2.168	2.312	2.272	2.056	1.945	2.150	E7	7.06
22)	Toxaphene-1							0.000		-1.00
23)	Toxaphene-2							0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD110618CLP.M
 Title : GC Extractables
 Last Update : Tue Nov 06 03:46:44 2018
 Response Via : Initial Calibration

Calibration Files

3 =PD050204.D 1 =PD050200.D 2 =PD050202.D
 4 =PD050206.D 5 =PD050208.D

	Compound	3	1	2	4	5	Avg	%RSD
24)	Toxaphene-3						0.000	-1.00
25)	Toxaphene-4						0.000	-1.00
26)	Toxaphene-5						0.000	-1.00
27) SA	Decachlorobiphenyl	1.824	2.067	1.984	1.675	1.545	1.819 E7	11.82

(#) = Out of Range