

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD110618\
 Data File : PD050207.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 05 Nov 2018 21:38
 Operator : AJ\SJ
 Sample : INDA501
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA501

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 06 03:09:02 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD110618CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Nov 06 03:08:52 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.233	4.060	50869680	2048.4E6	79.186	75.031
27)	SA Decachlor...	9.017	9.169	100.4E6	2612.8E6	154.344	146.861
Target Compounds							
2)	A alpha-BHC	4.686	4.453	76808783	3191.3E6	82.014	75.950
3)	MA gamma-BHC...	4.979	4.740	71916858	2886.1E6	80.507	75.795
4)	MA Heptachlor	5.291	5.254	70747239	2820.7E6	80.432	74.394
9)	A Endosulfan I	6.337	6.307	59992518	2138.7E6	79.797	74.350
13)	MA Dieldrin	6.578	6.566	140.8E6	4463.2E6	162.910	150.692
14)	MA Endrin	6.838	6.784	111.0E6	3481.9E6	160.781	151.820
16)	A 4,4'-DDD	6.928	6.902	107.1E6	3437.3E6	163.539	151.258
17)	MA 4,4'-DDT	7.169	7.202	111.4E6	3292.5E6	164.888	155.987
20)	A Methoxychlor	7.693	7.659	285.8E6	7036.0E6	790.587	743.315

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD110618\
Data File : PD050207.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 05 Nov 2018 21:38
Operator : AJ\SJ
Sample : INDA501
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
INDA501

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 06 03:09:02 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD110618CLP.M
Quant Title : GC Extractables
QLast Update : Tue Nov 06 03:08:52 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

