

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD110818\
Data File : PD050258.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 08 Nov 2018 21:55
Operator : AJ\SJ
Sample : J5844-01
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
MH-A-B

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 09 08:24:36 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD103018.M
Quant Title : GC Extractables
QLast Update : Tue Oct 30 03:08:01 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.234	4.060	11143177	461.4E6	17.775	16.566
28)	SA Decachlor...	9.017	9.169	11344526	311.0E6	18.872	18.095

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD110818\
Data File : PD050258.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 08 Nov 2018 21:55
Operator : AJ\SJ
Sample : J5844-01
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
MH-A-B

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 09 08:24:36 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD103018.M
Quant Title : GC Extractables
QLast Update : Tue Oct 30 03:08:01 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 4.234 min
Delta R.T.: 0.001 min
Response: 11143177
Conc: 17.77 ng/ml

Instrument :
ECD_D
ClientSampleId :
MH-A-B




#1 Tetrachloro-m-xylene

R.T.: 4.060 min
Delta R.T.: -0.001 min
Response: 461355910
Conc: 16.57 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.017 min
Delta R.T.: 0.000 min
Response: 11344526
Conc: 18.87 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.169 min
Delta R.T.: -0.004 min
Response: 311023076
Conc: 18.09 ng/ml