

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD110922\
 Data File : PD072900.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Nov 2022 16:09
 Operator : AR\AJ
 Sample : N5494-15
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 SB-22

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 11/10/2022
 Supervised By : Ankita Jodhani 11/10/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 09 21:12:22 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102822.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 28 15:29:46 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	2.701	3.462	29449528	577.3E6	14.359	16.726
2) SA Decachlor...	7.826	8.910	43039528	159.7E6	22.385m	16.664m#

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD110922\
Data File : PD072900.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Nov 2022 16:09
Operator : AR\AJ
Sample : N5494-15
Misc :
ALS Vial : 25 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
SB-22


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 11/10/2022
Supervised By : Ankita Jodhani 11/10/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 09 21:12:22 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD102822.M
Quant Title : GC Extractables
QLast Update : Fri Oct 28 15:29:46 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene

R.T.: 2.701 min
Delta R.T.: -0.003 min
Response: 29449528
Conc: 14.36 ng/ml

Instrument :
ECD_D
ClientSampleId :
SB-22


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 11/10/2022
Supervised By :Ankita Jodhani 11/10/2022




#1 Tetrachloro-m-xylene

R.T.: 3.462 min
Delta R.T.: -0.003 min
Response: 577347013
Conc: 16.73 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.826 min
Delta R.T.: -0.010 min
Response: 43039528
Conc: 22.39 ng/ml m



#28 Decachlorobiphenyl

R.T.: 8.910 min
Delta R.T.: -0.008 min
Response: 159682702
Conc: 16.66 ng/ml m