

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD110923\
 Data File : PD079504.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Nov 2023 12:14
 Operator : AR\AJ
 Sample : 05238-06
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 07 19:55:42 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD110623CLP.M
 Quant Title : GC Extractables
 QLast Update : Sat Nov 04 07:14:58 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
27)	SA Decachlor...	0.000	7.950	0	20222	N.D.	<MDL
Target Compounds							
4)	MA Heptachlor	0.000	3.969	0	106479	N.D.	<MDL
5)	MB Aldrin	0.000	4.240	0	248787	N.D.	<MDL
6)	B beta-BHC	0.000	3.969f	0	106479	N.D.	<MDL
7)	B delta-BHC	0.000	4.188f	0	194552	N.D.	<MDL
14)	MA Endrin	0.000	5.689	0	173690	N.D.	<MDL

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD110923\
Data File : PD079504.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Nov 2023 12:14
Operator : AR\AJ
Sample : 05238-06
Misc :
ALS Vial : 7 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 07 19:55:42 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD110623CLP.M
Quant Title : GC Extractables
QLast Update : Sat Nov 04 07:14:58 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm




Instrument :
ECD_D
ClientSampled :
0



#1 Tetrachloro-m-xylene


R.T.: 0.000 min
Exp R.T. : 3.567 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T. : 2.803 min
Response: 0
Conc: N.D.



#27 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 9.110 min
Response: 0
Conc: N.D.



#27 Decachlorobiphenyl

R.T.: 7.950 min
Delta R.T.: 0.000 min
Response: 20222
Conc: N.D.