

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD111023CLP-TCLP.M
 Title : GC Extractables
 Last Update : Thu Nov 09 02:57:06 2023
 Response Via : Initial Calibration

Calibration Files

3 =PD079514.D 1 =PD079512.D 2 =PD079513.D
 4 =PD079515.D 5 =PD079516.D 6 =PD079517.D

Compound	3	1	2	4	5	6	Avg	%RSD
1) SA Tetrachloro-m...	1.704	1.787	1.736	1.667	1.656	1.690	1.707 E6	2.84
2) A alpha-BHC							0.000	-1.00
3) MA gamma-BHC (Li...							0.000	-1.00
4) MA Heptachlor							0.000	-1.00
5) MB Aldrin							0.000	-1.00
6) B beta-BHC							0.000	-1.00
7) B delta-BHC							0.000	-1.00
8) B Heptachlor ep...							0.000	-1.00
9) A Endosulfan I							0.000	-1.00
10) B trans-Chlordane							0.000	-1.00
11) B cis-Chlordane							0.000	-1.00
12) B 4,4'-DDE							0.000	-1.00
13) MA Dieldrin							0.000	-1.00
14) MA Endrin							0.000	-1.00
15) B Endosulfan II							0.000	-1.00
16) A 4,4'-DDD							0.000	-1.00
17) MA 4,4'-DDT							0.000	-1.00
18) B Endrin aldehyde							0.000	-1.00
19) B Endosulfan Su...							0.000	-1.00
20) A Methoxychlor							0.000	-1.00
21) B Endrin ketone							0.000	-1.00
22) Toxaphene-1	1.537	1.661	1.570	1.504	1.568	1.555	1.566 E4	3.37
23) Toxaphene-2	1.876	1.936	1.889	1.853	1.891	1.963	1.901 E4	2.14
24) Toxaphene-3	5.345	5.288	5.349	5.483	5.628	5.901	5.499 E4	4.22
25) Toxaphene-4	4.010	3.929	3.960	4.093	4.277	4.493	4.127 E4	5.29
26) Toxaphene-5	2.977	2.839	2.989	3.110	3.190	3.295	3.067 E4	5.35
27) SA Decachlorobip...	2.018	2.064	2.050	1.964	1.940	1.992	2.005 E6	2.41

Signal #2 Calibration Files

3 =PD079514.D 1 =PD079512.D 2 =PD079513.D
 4 =PD079515.D 5 =PD079516.D 6 =PD079517.D

Compound	3	1	2	4	5	6	Avg	%RSD
1) SA Tetrachloro-m...	2.645	2.711	2.680	2.621	2.612	2.641	2.651 E6	1.41
2) A alpha-BHC							0.000	-1.00
3) MA gamma-BHC (Li...							0.000	-1.00
4) MA Heptachlor							0.000	-1.00
5) MB Aldrin							0.000	-1.00
6) B beta-BHC							0.000	-1.00
7) B delta-BHC							0.000	-1.00
8) B Heptachlor ep...							0.000	-1.00
9) A Endosulfan I							0.000	-1.00
10) B trans-Chlordane							0.000	-1.00
11) B cis-Chlordane							0.000	-1.00
12) B 4,4'-DDE							0.000	-1.00
13) MA Dieldrin							0.000	-1.00
14) MA Endrin							0.000	-1.00
15) B Endosulfan II							0.000	-1.00
16) A 4,4'-DDD							0.000	-1.00
17) MA 4,4'-DDT							0.000	-1.00
18) B Endrin aldehyde							0.000	-1.00
19) B Endosulfan Su...							0.000	-1.00
20) A Methoxychlor							0.000	-1.00
21) B Endrin ketone							0.000	-1.00
22) Toxaphene-1	1.636	1.621	1.652	1.617	1.597	1.623	1.624 E4	1.15

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD111023CLP-TCLP.M
 Title : GC Extractables
 Last Update : Thu Nov 09 02:57:06 2023
 Response Via : Initial Calibration

Calibration Files

3	=PD079514.D	1	=PD079512.D	2	=PD079513.D
4	=PD079515.D	5	=PD079516.D	6	=PD079517.D

	Compound	3	1	2	4	5	6	Avg		%RSD
23)	Toxaphene-2	1.668	1.785	1.688	1.663	1.663	1.680	1.691	E4	2.77
24)	Toxaphene-3	5.501	5.074	5.243	5.858	6.162	6.421	5.710	E4	9.26
25)	Toxaphene-4	7.622	7.213	7.304	7.808	7.852	7.982	7.630	E4	4.09
26)	Toxaphene-5	3.202	3.055	3.120	3.337	3.498	3.586	3.300	E4	6.42
27) SA	Decachlorobip...	2.642	2.768	2.706	2.586	2.565	2.582	2.642	E6	3.06

(#) = Out of Range